

Cite this: *Mater. Adv.*, 2025,
6, 117

Unlocking the potential of 2D nanomaterial-based biosensors in biomarker-based detection of *Helicobacter pylori*

Desmond Lutomia,^a Rachna Poria,^a Deepak Kala,^b Anupam Kumar Singh,^c
Manoj K Gupta,^d Deepak Kumar,^e  *^e Ankur Kaushal*^a and Shagun Gupta*^a

Recent advancements in nanotechnology and biomedicine have promoted the utilization of nanomaterials for various medical applications, particularly in the detection of *Helicobacter pylori* infections. The colonization of the gastric mucosa by *H. pylori* significantly establishes the risk factors for the development of chronic gastritis, peptic ulcers, and gastric cancers. While conventional methods, both invasive and non-invasive, are available for the detection of *H. pylori*, they often face limitations in terms of sensitivity, specificity, cost-effectiveness, and point-of-care applications. Two-dimensional nanomaterials exhibit considerable potential in the development of robust analytical platforms tailored for point-of-care (POC) detection of *H. pylori*, thereby presenting streamlined and economically viable biosensing solutions for the purpose of detecting *H. pylori* infections. This review summarizes the primary biomarkers utilized for the detection of *H. pylori* infections, elucidating how 2D nanomaterials enhance biosensor efficacy and the diverse applications of biosensors coupled with 2D nanomaterials in detecting *H. pylori* infection. Additionally, it examines different types of biosensing platforms that harness the unique properties of 2D nanomaterials and design considerations for optimizing biosensor performance for the accurate and reliable identification of *H. pylori* infections. The last part explores the industrialization potential and commercial viability and challenges inherent in utilizing 2D nanomaterials in biosensing and outlines future research and development prospects in 2D nanomaterial-based biosensors for disease detection.

Received 28th May 2024,
Accepted 1st November 2024

DOI: 10.1039/d4ma00546e

rsc.li/materials-advances

1. Introduction

H. pylori is a Gram-negative bacterium that colonizes the stomach in humans. This bacterium has adapted to survive in high acidic conditions of the stomach, where it colonizes the gastric mucosa, particularly the antrum and corpus regions and exists for most of its lifetime.¹ Morphologically, *H. pylori* has a curved or spiral rod shape of 2.5–5.5 μm in length and 0.5–1.5 μm in width.^{2,3} It has 2–6 unipolar, sheathed lophotrichous flagella

that enable mobility and chemotaxis within the viscous mucosal layer of the stomach.⁴ These flagella also aid adherence to gastric epithelial cells. *H. pylori* is microaerophilic, requiring 3–5% oxygen, 5–10% CO_2 , 0–10% H_2 , and high humidity to grow optimally.⁵ The bacterium possesses a unique urease enzyme that allows it to hydrolyze urea, producing ammonia and bicarbonate, which neutralize the acidic environment of the stomach and create a microenvironment favorable for bacterial survival. It also produces catalase and oxidase enzymes, which are used as biochemical factors for identification. The bacterial cell envelope consists of a cytoplasmic membrane, cell wall, and outer membrane containing lipopolysaccharides and membrane proteins. The helical cell shape, flagella, and outer membrane components are key virulence factors that promote colonization persistence in the hostile acidic stomach.

This bacterium colonizes the stomach of approximately 50% of the population globally, making it one of the most ubiquitous infections worldwide.⁵ However, prevalence rates vary geographically with higher rates noted in developing countries compared to industrialized nations. In Africa, *H. pylori* infects 60–90% of adults whereas prevalence is around 20–50%

^a Department of Bio-sciences and Technology, MM Engineering College, Maharishi Markandeshwar (Deemed to Be) University, Mullana, Ambala 134003, India.
E-mail: shagun.gupta@mmumullana.org, ankur.kaushal@mmumullana.org

^b NL-11 Centera Tetrahertz Laboratory, Institute of High Pressure Physics, Polish Academy of Sciences, 29/37 Sokolowska Street, 01142 Warsaw, Poland

^c Department of Gastroenterology, Postgraduate Institute of Medical Education and Research (PGIMER), Chandigarh, India

^d Department of Chemistry, School of Basic Sciences, Central University of Haryana, Mahendargarh-123031, Haryana, India

^e Department of Pharmaceutical Chemistry, School of Pharmaceutical Sciences, Shoolini University, Solan, Himachal Pradesh, 173229, India.
E-mail: guptadeepak002@gmail.com



in Western Europe and North America. In Asia, prevalence ranges from 22% in Japan to 87% in Bangladesh. India has an estimated *H. pylori* prevalence of 50–80% in adults.^{6,7} Low socioeconomic status and crowded living conditions correlate with higher infection risk. Primary transmission occurs through the fecal-oral and gastric-oral routes through contaminated food/water sources or close person-to-person contact. Infection is usually acquired in early childhood and without treatment, colonization persists lifelong. Given its high global prevalence, *H. pylori* infection is believed to play a causal role in 75–80% of gastric cancer cases worldwide.^{8,9} Epidemiologic patterns correspond with gastric cancer incidence in both developing and developed regions. Eradication of *H. pylori* can lead to regression of gastric precancerous lesions and reduce cancer risk. Hence many health organizations recommend routine testing and treatment to prevent long-term complications, especially in high-risk populations. However, reinfection after antibiotic therapy remains a challenge especially in developing countries due to continued exposure.

H. pylori colonization triggers chronic active gastritis in all infected persons but only 10–20% develop more severe complications like ulcers and cancer.^{8–10} Colonization initiates persistent inflammation through bacterial virulence factors and dysregulated host immune responses. Gastric epithelial cell injury arises due to direct cytotoxicity mediated by factors like the vacuolating cytotoxin A (*VacA*), the oncoprotein cytotoxin-associated gene A (*CagA*), and urease-driven pH alterations.^{8,11} Robust pro-inflammatory Th1-cell mediated responses further damage the mucosa.¹¹ Concurrently, *H. pylori* evades immune clearance through myriad mechanisms thereby establishing persistent infection. The pattern and severity of gastritis and epithelial damage determines specific disease outcomes. Antral-predominant colonization increases acid secretion and causes duodenal ulcers. Body-predominant gastritis reduces acid levels leading to atrophic gastritis, a precursor of gastric cancer. Surface mucosal damage allows back-diffusion of acid that worsens inflammation.¹¹ Ulcers arise when this exceeds the tissue's reparative capacity. Atrophic gastritis accompanied by mutagenic properties of *H. pylori* promotes oncogenic mutations in gastric stem cells over decades.^{12,13} The stepwise progression from chronic gastritis through precancerous lesions culminates eventually in gastric adenocarcinoma. Gastric mucosa-associated lymphoid tissue (MALT) lymphoma is another serious complication attributed to chronic *H. pylori*-triggered immunostimulation.¹⁴ Despite its prevalence, not all individuals infected with *H. pylori* develop symptomatic disease, with factors such as bacterial virulence factors, host genetics, and environmental factors playing a role in disease manifestation.

Early diagnosis and eradication of *H. pylori* infection is crucial to prevent progression to severe gastroduodenal complications, especially in high-risk populations allowing for timely intervention and treatment. Furthermore, early detection can help reduce the risk of transmission to others and minimize the burden of disease in affected individuals and healthcare systems. However, detection of *H. pylori* can be challenging as up to 50–80% of infected individuals are asymptomatic.^{7,15} For the

detection of *H. pylori*, two approaches – invasive and non-invasive techniques – are utilized. Fig. 1 illustrates invasive, noninvasive, and biosensor techniques used for the diagnosis and detection of *H. pylori* infection. Conventional invasive tests to detect active infection require endoscopy to obtain gastric mucosal biopsies followed by culture, histology, rapid urease testing, or PCR.¹⁵ Endoscopy allows direct visualization of the stomach lining but is uncomfortable for patients, has procedure-related risks, and is expensive. Non-invasive indirect tests like the urea breath test (UBT), stool antigen test, and serology are easier and cheaper but have reduced accuracy. ¹³C-UBT and ¹⁴C-UBT are widely used for initial diagnosis and post-treatment confirmation; patients ingest ¹³C- or ¹⁴C-labeled urea which gets hydrolyzed by *H. pylori*'s urease, releasing labeled CO₂ detected in breath samples.¹⁶ Stool antigen ELISA detects *H. pylori* proteins in fecal samples.¹⁶ Both methods have high sensitivity and specificity but cannot gauge severity or locate lesions. Serology detects serum antibodies to *H. pylori* but remains positive even after clearance, limiting use for test-of-cure. Sensitivity also varies based on local infection prevalence affecting the positive predictive value. The constraints of these conventional tests include the requirement for reliably cultured bacteria, lack of quantitation, and inability to evaluate pathological changes or antibiotic susceptibility. Rapid molecular techniques like fluorescence *in situ* hybridization (FISH) overcome some limitations due to direct detection of *H. pylori* DNA/RNA in tissues or body fluids.¹⁷ But invasive sampling restricts routine applicability. Non-invasive and highly sensitive urine, saliva or blood PCR-based assays are being developed but remain experimental. Thus, while a range of methods are available for detecting active *H. pylori* infection, no single test meets all clinical needs regarding speed, cost, accuracy, and ability to monitor pathogenesis or response to therapy. More convenient and quantitative point-of-care tests are needed, especially in resource-limited settings with high disease burden.

The limitations of conventional techniques have driven research into nanomaterial-based biosensors for improved *H. pylori* detection (Table 1). Biosensors offer a non-invasive, rapid, and cost-effective alternative for *H. pylori* detection, enabling point-of-care testing and improving access to timely diagnosis for patients.

These biosensors incorporate a biological sensing element with a transducer to yield a detectable signal proportional to the target pathogen. 2D nanomaterials exhibit diverse properties in terms of their mechanical, chemical, and optical attributes, apart from uniqueness in their size, aspect ratio, shape, biocompatibility, and biodegradability. These varied properties afford a suitable platform for these nanomaterials for myriad applications in the healthcare sector, including drug delivery, bioimaging, tissue engineering, biosensors, and so on. However, their low dimensionality and exceptionally large aspect ratios (surface area-to-volume ratio) make these nanomaterials invaluable for uses demanding high levels of surface interactions in the nanoscale regime.^{33,34} Specifically, 2D nanoparticles like graphene oxide (GO) allow for efficient immobilization of biomolecules, such as antibodies or nucleic acid-based probes, to capture *H. pylori* cells or antigens through strong π - π stacking





Fig. 1 Invasive, noninvasive, and biosensor techniques used for the diagnosis and detection of *H. pylori* infection.

interactions, enhancing the specificity of detection.³⁵ This review summarizes the overview of *H. pylori* infection and associated diseases, possible biomarkers of *H. pylori*, and how integration of 2D nanoparticles in biosensors enhances the detection of *H. pylori* infection. Challenges associated with biosensing technology for diagnostic purposes, recent advancements and future perspectives in the detection of *H. pylori* infection are covered in the last section.

2. Biomarkers associated with *H. pylori* infection

H. pylori utilizes a variety of virulence factors and mechanisms to induce pathogenesis while evading host immunity.⁴ These include both bacterial components and products that directly damage gastric tissues as well as factors that help in its persistence and proliferation.^{5,6} Some key general virulence attributes are the spiral shape, flagella, and urease production.^{4,10,11,36}

The helical cell morphology along with flagella enables motility and chemotaxis allowing it to penetrate and survive within the viscous mucosal layer.⁸ The outer membrane lipopolysaccharides (LPS) and other surface molecules help evade phagocytosis and complement-mediated killing.^{8,11} The urease enzyme is a vital virulence factor for colonization as it catalyzes urea breakdown into ammonia and CO₂, neutralizing gastric acid and forming a microenvironment conducive for bacterial growth. Urease is encoded by the *ureA*, *ureB*, and *ureC* genes and expression of the accessory genes *ureI*, *ureE*, *ureF*, *ureG*, and *ureH* results in the catalytically active urease enzyme.^{37,38} Due to the high specificity and sensitivity of the urease enzyme, it has been shown to be a significant biomarker for *H. pylori* infection detection.³⁶ Studies have shown that urease activity correlates with the severity of gastritis and peptic ulcer disease (PUD), making it a valuable biomarker for disease monitoring and treatment response assessment.³⁹

H. pylori secretes two major cytotoxins (*VacA* and *CagA*) which are the major virulence factors as well as biomarkers

Table 1 Performance metrics of commonly used conventional diagnostic tests for *H. pylori* detection

Current methods	Limitation	Sensitivity (%)	Specificity (%)	Time	Ref.
Endoscopy	Uncomfortable, highly expensive, risk of infection, risk of bleeding	95	99	7 days	18–20
Histology	Sample error (sampling can lead to false negatives)	95	99	7–10 days	21
Urea breath test	Recent use of PPIs or antibiotics reduces the bacterial load	90	95	1 h	16, 22 and 23
Culture	More time consuming	58	99	7 to 10 days	20, 24 and 25
Urease test	Recent use of PPIs or antibiotics reduces the bacterial load	95.2	95.1	1 day	26–28
Stool antigen test	If the sample is not stored properly, the antigen degrades over time	95	96	2–4 days	29–31
Serology	Cannot differentiate between past and recent infection	75–85	79–90	More than 3 h	20, 29 and 32



associated specifically with *H. pylori* infection. The toxin vacuolating cytotoxin A (*VacA*) induces cellular vacuolization and injury.³⁹ It also increases permeability of epithelial layers enabling *H. pylori* access to underlying tissues. Studies have shown that the *VacA* S1 genotype is highly associated with increased levels of interleukin-8, chronic inflammation in gastric mucosa, epithelial damage and metaplasia as compared to the *VacA* S2 genotype.^{40,41} The cytotoxin-associated gene A (*CagA*) protein is among the most extensively studied and occurs in 60–70% of isolates globally. It is located in the *cag* pathogenicity island (*cagPAI*), a 40 kb DNA segment encoding a type IV secretion system that forms needle-like pili.^{42,43} The *CagA* effector protein is injected *via* this secretion system directly into host cells where it dysregulates signaling through intramolecular interaction and causes cellular defects.⁴² *CagA* by interacting with multiple host signaling molecules such as the pro-oncogenic phosphatase (*SHP2*) and the polarity-regulating kinase (*PAR1*) induces pro-inflammatory responses by activating NF- κ B in a phosphorylation-dependent manner.⁴² Fig. 2 describes the major pathogenic markers of *H. pylori* and the delivery of the *CagA* gene into the epithelial cell through the T4SS and the intramolecular interaction disrupting cellular signaling.^{41,42} Recent research has focused on elucidating the mechanisms underlying *CagA*-mediated pathogenesis and its role as a biomarker for disease risk stratification and prognosis assessment.^{42,44} These studies show that *CagA*- and

cagPAI-positive strains are associated with heightened inflammation and increased risk for severe gastritis, peptic ulcers and cancer compared to *cagPAI*-negative isolates. A meta-analysis has associated *CagA*-positive strains with a 2-fold increased risk for peptic ulcer disease and 5-fold higher risk for gastric cancer compared to *cagPAI*-negative isolates.⁴⁴ *CagA* also causes cellular changes that may promote carcinogenesis including disruption of cell-to-cell adhesion, elongation of epithelial cells, and abnormal proliferation. Another *cagPAI* effector called *CagL* binds to and activates β 1 integrin receptors on gastric cells triggering pro-inflammatory signaling.⁴⁵ The *cagPAI* genes also induce the production of cytokines like interleukin-8 by gastric epithelial cells which drive neutrophil infiltration and chronic inflammation.^{42,44} *CagA* is the first bacterial oncoprotein to be identified and remains a major virulence factor driving *H. pylori*-mediated injury and carcinogenesis through its multifaceted effects on gastric epithelial cells.¹¹ Furthermore, studies have explored the use of *CagA*-specific antibodies and DNA probes in biosensor-based diagnostic assays for the detection of *H. pylori* infection, highlighting the potential clinical utility of *CagA* as a diagnostic biomarker.⁴⁶

Other ubiquitous virulence determinants that may serve as biomarkers for pathogenic potential are adhesins like *Baba*, *Saba*, *pylori AlpA*, *AlpB* and *HopZ* that mediate binding to gastric epithelial cells, an essential initial step in colonization. The blood group antigen binding adhesin (*Baba*) is encoded by the *Baba2* gene, and shows increased association with



Fig. 2 The major pathogenic markers of *H. pylori* and the delivery of the *CagA* gene into the epithelial cell through the T4SS and the intramolecular interaction disrupting cellular signaling.



duodenal ulcer (71–83%) and gastric cancer (83–85%) compared to gastritis alone (22–59%).^{11,15,39,47} The outer inflammatory protein A (*OipA*) induces interleukin-8 secretion and neutrophil infiltration.⁴⁷ *OipA* is linked to higher inflammation severity. The *dupA* gene stimulates IL-12 and is variably associated with duodenal ulcers and reduced risk of gastric cancer.⁴⁷ Another biomarker gene is the *H. pylori* cholesterol- α -glucosyltransferase (*CGT*) gene which is involved in the synthesis of lipopolysaccharides of the cell wall. The *CGT* catalyzes the conversion of membrane cholesterol to cholesteryl glucosidases which are incorporated in the cell wall enabling the bacteria to evade phagocytosis by the immune cells.⁴⁷ The catalase enzyme is also secreted and plays a role in protecting the bacteria against oxidative damage by breaking down harmful reactive oxygen species. Superoxide dismutase and alkyl hydroperoxide reductase further enhance resistance to oxygen radicals.⁴⁷ The arginase enzyme is also secreted and allows *H. pylori* to compete with host cells for L-arginine, thereby impairing T-cell function.⁴⁷ Other emerging biomarkers include *iceA* alleles, homologues of the helicobacter outer membrane *PorD* protein, and *fur* mutations which are postulated to enhance *H. pylori* virulence but require more confirmatory studies.^{48,49} Host genetic factors like IL-1 β polymorphisms combined with specific bacterial alleles are also increasingly investigated as joint biomarkers.⁴⁸ Many studies have provided insights into the molecular mechanisms underlying the actions of these virulence factors and their potential as diagnostic and therapeutic targets for *H. pylori*-associated diseases. Recent investigations have revealed that a strategic combination of biomarkers significantly enhances the diagnostic precision for *H. pylori*-associated gastric and duodenal cancers. By employing isobaric tag for relative and absolute quantitation, researchers have analyzed protein expression levels in patients infected with *H. pylori*. The optimized integration of virulence factors *OipA*, *Baba*, and *Saba* achieved a predictive accuracy of 77.3% in differentiating between gastric cancer and duodenal ulcer patients.⁵⁰ This combination not only improves diagnostic precision but also suggests that a protein microarray incorporating these antigens could facilitate rapid and efficient diagnosis of *H. pylori*-related gastric cancer. Moreover, the incorporation of these biomarkers with advanced technologies such as 2D nanomaterials underscores their potential in developing rapid diagnostic sensors that enable early detection and timely intervention.

3. Integration of 2D nanoparticles in biosensors for detection of *H. pylori*

3.1. Properties and synthesis of 2D nanoparticles

Nanomaterials such as graphene, graphene oxide, hexagonal boron nitride, transition metal dichalcogenides (TMDs), MXenes, layered double hydroxides, transition metal oxides, and black phosphorus have emerged as promising 2D nanomaterials for biosensing platforms. This is attributed to their exotic structural, electrical, optical, and catalytic properties originating from the quantum confinement in 2D layers with

atomic-scale thickness.^{33,34,51,52} For instance, graphene exhibits exceptionally high electron mobility, mechanical strength, specific surface area and thermal conductivity. Its large surface area allows for efficient immobilization of biomolecules, such as antibodies or DNA probes, enhancing the sensitivity and specificity of biosensors. Moreover, the high conductivity of graphene enables rapid electron transfer, resulting in fast and accurate detection of analytes. Recent literature has demonstrated the integration of graphene-based biosensors for the detection of various biomarkers, including proteins, nucleic acids, and pathogens. For example, graphene-based biosensors have been utilized for the detection of cancer biomarkers, infectious diseases, and environmental pollutants, offering sensitive and selective detection platforms for diagnostic purposes.²⁰ The smart health watch “GF1” is one such graphene-based commercially available product launched by Wuxi Graphene Film (a subordinate of The Sixth Elements Materials). Chemical vapor deposited graphene film acts as a touch screen conductive element in lieu of commonly used indium tin oxide glass. GF1 is a smart health monitoring watch that provides a detailed health index and scrutinizes the dynamic electrocardiogram, heart rate, temperature, and blood pressure, apart from tracking the calories burned during physical workouts.³⁶ The integration of molybdenum disulfide (MoS₂) in biosensors allows for enhanced sensitivity and selectivity, as well as improved stability and biocompatibility. Studies have highlighted the potential of MoS₂-based biosensors for point-of-care diagnostics, environmental monitoring, and biomedical research. For instance, MoS₂-based biosensors have been utilized to detect glucose, cholesterol, and other metabolites in clinical samples, offering rapid and accurate measurements for disease diagnosis and management.⁵³ MXenes containing hydroxyl or fluoride terminations readily adsorb biomolecules electrostatically. Layered black phosphorus demonstrates superior anisotropic electrical properties, chemical inertness, thermal stability, and a wide bandgap optimal for field-effect biosensing. Such unique characteristics make 2D nanomaterials extremely attractive for fabricating electrochemical, optical, piezoelectric or field-effect biosensors with high sensitivity and selectivity.

2D nanoparticles are generally synthesized through three broad approaches – top-down approaches, bottom-up synthesis, and hybrid methods.^{54,55} The top-down approach involves splitting bulk crystals into nanosheets using mechanical or liquid phase exfoliation. Mechanical exfoliation repeatedly peels off layers using adhesive tape but has very low yield. Liquid exfoliation uses ultrasonication or chemical intercalation in solvents like *N*-methyl-2-pyrrolidone (NMP) to produce dispersions of 2D flakes. This is simple and scalable but can damage sheets and provide limited control over thickness. The bottom-up synthesis builds 2D layers atom-by-atom *via* chemical vapor deposition, solvothermal synthesis, or self-assembly. While high-quality sheets are produced, these methods require high temperatures and complex setups.^{56–59} Hybrid techniques combine top-down and bottom-up methods by exfoliating bulk crystals and then using the nanosheets as 2D templates for further synthesis through chemical vapor deposition (CVD) or atomic



layer deposition. Each approach yields nanomaterials suited for particular biosensing modalities. For instance, liquid exfoliated graphene oxide (GO) with its abundant surface oxygen groups has been widely used to develop electrochemical immunosensors and fluorescent DNA biosensors for diagnostic purposes.⁶⁰ CVD-synthesized MoS₂ nanosheets are integrated in field-effect transistors due to their semiconductor properties. Mechanically exfoliated boron nitride nanosheets are ideal for surface-enhanced Raman spectroscopy based detection. Thus, based on the desired transduction mechanism, suitable 2D nanomaterials can be synthesized using an optimal technique and integrated to develop highly efficient biosensors for clinical applications and diagnosis of *H. pylori* infection.

3.2. Surface functionalization of 2D nanoparticles for biosensing applications

Efficient functionalization and modification facilitate immobilization of biorecognition elements and improve interaction with target analytes. The utilization of 2D nanomaterials in biosensing applications transcends conventional detection methods, encompassing capabilities for real-time monitoring and point-of-care diagnostics. The high surface area and ease of functionalization of these nanomaterials significantly enhance their performance in dynamic biological environments, making them particularly suitable for continuous health monitoring applications.⁶¹ This adaptability allows for the development of biosensors that can operate effectively in complex biological matrices, thereby improving diagnostic accuracy and reliability. The clinical implications of employing functionalized 2D nanoparticles in biosensing are substantial, as they have the potential to revolutionize diagnostic practices. By facilitating early detection of diseases through enhanced sensitivity and specificity, these advanced biosensors can lead to improved patient outcomes.^{62,63} Moreover, the integration of 2D nanomaterials into biosensing platforms supports the implementation of personalized healthcare solutions, enabling clinicians to tailor treatments based on individual patient profiles.

Two main common approaches for functionalizing 2D nanoparticles involve covalent and non-covalent strategies.^{64–66} Covalent approaches involve covalent attachment of biomolecules, such as antibodies, aptamers, or DNA probes, onto the nanoparticle surface forming stable bonds between surface groups like –COOH, –NH₂, and –OH and biomolecules *via* crosslinkers, providing precise control over orientation and conformation. Non-covalent methods utilize electrostatic, π – π stacking or hydrophobic interactions for bioconjugation. These are simpler, preserve nanomaterial properties better but provide less control over binding. For example, graphene oxide contains abundant surface carboxyl and hydroxyl groups amenable to covalent tethering of enzymes, antibodies or aptamers using EDC/NHS chemistry.⁶⁷ TMD nanosheets can be functionalized by silanization of surface –OH groups, allowing covalent protein attachment. MXenes with –OH, –O and –F terminations readily adsorb biomolecules non-covalently through electrostatic and hydrogen bonding. Proper bioconjugation and blockage of unused sites is vital for optimal biorecognition while minimizing non-specific binding.

The literature has shown that suitable chemical or physical functionalization facilitates the immobilization of bioreceptors on 2D surfaces and enables specific, oriented binding of analytes in biosensing applications.^{68–72}

In addition to biomolecule functionalization, surface engineering strategies have been employed to modify the properties of 2D nanoparticles for biosensing applications. One such strategy involves modifying the surfaces of 2D nanoparticles with metal nanoparticles, such as gold nanoparticles (AuNPs) or silver nanoparticles (AgNPs), to enhance their sensing capabilities. Metal-decorated 2D nanoparticles exhibit localized surface plasmon resonance (LSPR) effects, which can amplify the signal response upon target binding, leading to improved detection sensitivity.⁷³ For instance, AuNP-modified graphene-based biosensors have been developed for the ultrasensitive detection of SARS-CoV-2 antigens in clinical samples.⁷⁴ Another surface modification approach involves the introduction of defect sites or functional groups onto the nanoparticle surface to create specific binding sites for target molecules. Functionalized defect-engineered graphene has been used to selectively capture DNA sequences in biological samples, enabling the rapid and sensitive detection of the infectious pathogens in clinical samples.⁷⁴ The graphene biosensor based on a field-effect transistor (FET) architecture, functionalized by phage tail spike proteins (TSPs), has been shown to accurately detect pathogens such as *E. coli* at the single bacterium level.^{75,76} Additionally, a DNA-functionalized graphene FET has been proposed for the quantitation of proteins, utilizing the ability of DNA hybridization to amplify the detection signal.⁷⁷ Graphene nanosheets modified on a glassy carbon electrode have demonstrated enhanced electron transfer ability and interface adsorption capacity, leading to improved sensitivity in the detection of rutin, nitrofurazone and linagliptin.^{78,79} Furthermore, an armchair graphene nanoribbon (AGNR) interconnected between gold electrodes has been investigated as a detector of DNA hybridization, showing high sensitivity and accuracy in detecting changes in electrical properties upon functionalization with probe and target DNA.^{80,81} These surface engineering strategies offer versatile means of tailoring the properties of 2D nanoparticles for biosensing applications, allowing for the development of highly sensitive and selective biosensors for analyte detection.

Another approach involves the incorporation of 2D nanoparticles in microfluidic systems, where target molecules are selectively captured and concentrated within microfluidic channels, enabling sensitive and rapid detection. Microfluidic biosensors based on MoS₂ have been developed for the immune detection of *Toxoplasma gondii* in serum, offering a portable and point-of-care diagnostic platform.⁸² The biosensor showed a detection range of 1 pg mL⁻¹ to 10 ng mL⁻¹ for *T. gondii* monoclonal antibody solutions with a sensitivity of 3.358 nm/log (mg mL⁻¹). The biosensor demonstrated excellent specificity and clinical characteristics when tested with rabies virus, pseudorabies virus, and *T. gondii* serum, indicating its potential for biomedical applications.⁸² Integration of biosensing platforms leverages the unique properties of 2D nanoparticles to achieve



sensitive, selective, and rapid detection of infectious pathogens such as *H. pylori* biomarkers, paving the way for the development of next-generation POC diagnostic devices for the early diagnosis and monitoring of *H. pylori*-related diseases.

3.3. Biosensing platforms for *H. pylori* detection

Biosensors are qualitative and semiquantitative analytical devices that translate a biological recognition reaction into a measurable signal. They integrate biological components like enzymes, antibodies, nucleic acid, and cells with a physicochemical transducer.⁸³ Upon interaction with the analyte, the biorecognition element undergoes a specific biochemical reaction, leading to a change in the physicochemical properties of the transducer, which is then converted into an electrical, optical, or mechanical signal. This signal is subsequently amplified, processed, and displayed by a readout system, allowing for the quantitative analysis of the target analyte. These features have revolutionized biosensors as the point-of-care detection technique for various infections by offering rapid, sensitive, and specific diagnostic capabilities. The development of advanced biosensing platforms for the detection of *H. pylori* infection represents a crucial bridge between scientific innovation and clinical practice. Multiplexing capabilities of these platforms significantly enhance diagnostic precision by allowing simultaneous detection of multiple biomarkers, offering a more detailed analysis of a patient's condition.⁵⁰ This approach facilitates the development of personalized therapeutic strategies tailored to individual biomarker profiles. Clinically, these biosensing technologies hold promise for delivering rapid and highly reliable diagnostics, leading to more effective patient management and improved health outcomes.^{84–86} Furthermore, their adaptability for point-of-care use positions these biosensors as invaluable tools in resource-constrained environments where conventional diagnostic infrastructure may be unavailable.^{20,87}

Recent studies revealed that several electrochemical, optical, field-effect transistor and piezoelectric biosensors have been developed for point-of-care detection of *H. pylori* infection. Immunosensors utilizing *H. pylori* antibodies as biorecognition elements are integrated into formats like enzyme-linked immunosorbent assay (ELISA), chemiluminescence, and fluorescence platforms. DNA biosensors incorporating ssDNA probes complementary to *H. pylori* 16S rRNA or 23S rRNA provide rapid hybridization-based analysis. Aptasensors using synthetic aptamers selected against *H. pylori* whole cells or specific antigens offer high selectivity. Emerging 2D nanomaterials have significantly improved the performance of biosensors for *H. pylori* detection. Recent literature has evaluated the performance of different biosensor systems in terms of sensitivity, specificity, selectivity, and limit of detection. These 2D nanoparticle-based biosensors demonstrated high selectivity, good stability, reproducibility, and cost-effectiveness for early detection of *H. pylori*.^{87–95} These biosensors offer superior performance compared to conventional methods such as enzyme-linked immunosorbent assay (ELISA) and polymerase chain reaction (PCR), which typically have higher LODs and longer assay times.

Table 2 shows a comparison of performance parameters among 2D nanoparticle-based biosensors, biosensors with other nanomaterials, and conventional diagnosis methods. 2D nanoparticles have been integrated into optical, colorimetric and electrochemical biosensing platforms as well as field-effect transistors and piezoelectric sensors due to their unique properties and ease of functionalization.^{27,85,86,96} The integration of 2D nanoparticles into these sensing platforms has further advanced their detection capabilities by enhancing sensitivity, specificity, and selectivity.

3.3.1. Electrochemical-based biosensors. Electrochemical biosensors have emerged as promising tools for the rapid, accurate, and timely detection of *H. pylori* infection. Researchers have been actively exploring various approaches to improve the sensitivity, specificity, and convenience of electrochemical biosensors for *H. pylori* diagnosis. Electrochemical biosensors provide an attractive means to analyze the content of a biological sample due to the direct conversion of a biological event to an electronic signal. The signal transduction and the general performance of electrochemical sensors are often determined by the surface architectures that connect the sensing element to the biological sample at the nanometer scale. The most common surface modification techniques, the various electrochemical transduction mechanisms, and the choice of the recognition receptor molecules all influence the ultimate sensitivity of the sensor.⁹⁶ Fig. 3 describes the basic principle and working mechanism of the electrochemical biosensors for *H. pylori* detection. To improve the functionality and stability of these sensors, researchers have developed a method of integrating 2D nanoparticles enabling the reliability of these sensors for use in biomedical applications. Jaradat *et al.* developed an electrochemical immunosensor for detecting the *HopQ* protein, a biomarker present in saliva samples for the detection of *H. pylori*.⁸⁷ Consequently, this immunosensor exhibited high sensitivity and excellent linearity within the range of 10 pg mL⁻¹ to 100 ng mL⁻¹, with a remarkably low limit of detection of 10 pg mL⁻¹. Moreover, it demonstrated desirable characteristics such as good stability, reproducibility, and cost-effectiveness, rendering it a promising tool for early detection of *H. pylori*.⁸⁷

In a recent study, Mirzaei *et al.* introduced a novel Hsp60 biosensor with enhanced sensitivity and selectivity, developed using a nanocomposite with an aptamer-recognition surface for enhancing *H. pylori* detection.¹⁴⁶ The biosensor was constructed by modifying a glassy carbon electrode with a nanocomposite composed of carbon quantum dots, gold nanoparticles, and polythiophene. The integration of these materials provided the biosensor with high stability, a large surface area, excellent electrical conductivity, and compatibility with biological systems. Electrochemical techniques, including square wave voltammetry (SWV) and electrochemical impedance spectroscopy (EIS), were used to detect the interaction of Hsp60 with immobilized aptamers, resulting in a decrease in peak current and increased resistance. The aptasensor demonstrated a detection limit of 7.38 nM and a linear detection range of 0.01 to 0.25 μ M.¹⁴⁶ It also exhibited high reproducibility, selectivity, rapid response, and stability for detecting *H. pylori*,





Table 2 Comparison of performance parameters among 2D nanoparticle-based biosensors and other conventional diagnosis methods

Features	2D Nanoparticle-based biosensors	Biosensors with other nanomaterials	Conventional diagnosis methods (RUT, UBT, culture, endoscopy, PCR)	Ref.
Fabrication and integration processes	Complex fabrication, specialized equipment and expertise required. Integration involves functionalization of 2D nanoparticles with recognition elements with recognition elements	Fabrication may vary based on nanomaterials used, involving specialized equipment. Integration requires functionalization of nanomaterials with recognition elements	Standardized procedures, requiring specialized equipment and trained personnel	29, 64 and 66
Stability	Moderate stability, susceptible to degradation over time due to environmental factors and biofouling	Moderate stability, may experience degradation over time	Stable under appropriate conditions	69 and 97–99
Durability	Moderate durability, may degrade over prolonged use	Moderate durability, may degrade over prolonged use	Durable, can be used repeatedly	28 and 100
Scalability	Limited scalability due to complex fabrication processes and material availability	Moderate scalability, depending on nanomaterial synthesis and functionalization techniques	Scalable for mass production	101 and 102
Reproducibility	Variable reproducibility due to batch-to-batch variations in nanoparticle synthesis and functionalization methods	Variable reproducibility, influenced by nanomaterial synthesis and functionalization methods	High reproducibility, standardized procedures	103–105
Toxicity	Concerns regarding potential toxicity of 2D nanoparticles, necessitating thorough biocompatibility testing	Potential toxicity of other nanomaterials, necessitating biocompatibility assessment	Generally safe, but potential risks associated with invasive procedures	98 and 106–108
Biocompatibility	Requires thorough assessment of biocompatibility due to potential interactions with biological systems	Requires assessment of biocompatibility based on specific nanomaterials used	Biocompatible, suitable for use in biological systems	100, 101 and 109
Detection time	Rapid detection, often within minutes to hours	Rapid detection, similar to 2D nanoparticle-based biosensors	Variable detection time, depending on the method used: RUT (1–24 hours), UBT (30 minutes), culture (3–14 days), endoscopy (same day), PCR (3–12 hours)	87, 93, 101, 110 and 111
Limit of detection	Low limit of detection, down to picomolar or femtomolar levels	Low limit of detection, similar to 2D nanoparticle-based biosensors	Variable limit of detection, influenced by the sensitivity of the method used: RUT (qualitative), UBT (qualitative), culture (10^2 CFU mL^{-1}), endoscopy (qualitative), PCR (10^2 CFU mL^{-1})	24, 28, 29, 87, 88, 101, 103, 112 and 113
Selectivity	High selectivity, can distinguish the target analyte from interfering substances	Variable selectivity, influenced by nanomaterial properties and surface functionalization	High selectivity based on specific biomolecular interactions	90, 101, 103, 114 and 115
Sensitivity	High sensitivity, capable of detecting low concentrations of the target analyte (92–98%)	Variable sensitivity, influenced by nanomaterial properties and detection method (88–96%)	Variable sensitivity, depending on the method used: RUT (70–90%), UBT (90–98%), PCR (94–98%), ELISA (86, 101, 103 and 116–118)	9, 15, 17, 24, 86, 101, 103 and 116–118
Specificity	High specificity, able to accurately identify the target analyte amid complex biological matrices (94–99%)	Variable specificity influenced by nanomaterial properties and recognition elements (90–97%)	High specificity, based on specific biomolecular interactions: RUT (80–95%), UBT (92–98%), PCR (95–99%), ELISA (85–95%)	9, 14–17, 24, 36, 86, 99, 101, 103 and 116–118
Material and production cost	Moderate to high upfront costs, but cost-efficient in large-scale production due to scalable manufacturing techniques. These materials are typically cheaper than conventional ones once production is scaled up	Similar initial cost as the 2D nanomaterial-based biosensors	Typically high due to expensive reagents, specialized lab equipment, and complex infrastructure requirements	119–121



Table 2 (continued)

Features	2D Nanoparticle-based biosensors	Biosensors with other nanomaterials	Conventional diagnosis methods (RUT, UBT, culture, endoscopy, PCR)	Ref.
Design and maintenance cost	Compact and portable designs possible, with low to moderate cost of maintenance. These biosensors are often portable, disposable, or require minimal maintenance, reducing long-term upkeep	Similar design and cost considerations as 2D nanoparticle-based biosensors, with varying costs depending on the nanomaterial and fabrication processes	Standardized design and costs based on established protocols, with endoscopy and PCR being more expensive than RUT and UBT. High maintenance cost requiring regular calibration, expensive reagents, and highly trained personnel	17, 27, 98 and 121–124
Result consistency	Variable consistency due to batch-to-batch variations in nanoparticle synthesis and functionalization	Variable consistency, influenced by nanomaterial properties	Varying consistency: RUT (88–95%), UBT (90–95%), culture (90–98%), endoscopy (90–95%), PCR (95–99%)	28, 101, 114 and 115
Acceptability to the public	Moderate acceptability due to the novelty of nanomaterial-based biosensors and potential concerns about toxicity	Similar acceptability challenges as 2D nanoparticle-based biosensors, with varying public perception based on the nanomaterial and potential toxicity concerns	Conventional diagnostic methods are generally well-accepted by the public, with the exception of invasive procedures like endoscopy	17, 118, 125 and 126
Commercialization	Moderate commercialization potential, with several companies and startups exploring nanomaterial-based biosensors	Similar commercialization potential as 2D nanoparticle-based biosensors, with varying market readiness based on the nanomaterial and sensing mechanism	Widely commercialized and available in clinical settings, with varying degrees of accessibility and affordability	127–130
Real-world applications	Potential for rapid and decentralized diagnostics during disease outbreaks, with the ability to scale up production and deployment	Similar potential applications as 2D nanoparticle-based biosensors, with varying suitability based on the nanomaterial and sensing mechanism	Conventional diagnostic methods can be deployed during epidemics, but may face challenges in terms of accessibility, turnaround times, and logistics	29, 53, 89, 90, 100, 101, 103, 106, 115 and 127–145




Fig. 3 Electrochemical biosensors for *H. pylori* detection. When the target interacts with the immobilized bioreceptors, they cause biochemical reactions which are converted into measurable electrical signals, enabling the detection and quantification of the target.

highlighting its potential for biomedical, bioengineering, and clinical applications.

Saxena *et al.* conducted research on a molecularly imprinted polymer (MIP)-based electrochemical biosensor to detect the *VacA* antigen.⁸⁶ This was achieved by polymerizing the template *VacA* antigen on a paper electrode. Notably, the *VacA* antigen utilized MIP coated on a silicon dioxide (SiO_2) substrate integrated into a screen-printed paper electrode (SPPE) platform (*VacA*-MIP/ SiO_2 /SPPE sensor) exhibited exceptional sensitivity ($0.304 \text{ mA ng}^{-1} \text{ mL}^{-1}$) and an exceptionally low limit of detection (0.01 ng mL^{-1}) within a linear range of $0.01\text{--}100 \text{ ng mL}^{-1}$.⁸⁶ Furthermore, the sensor demonstrated successful determination of the *VacA* antigen in the presence of various potential interferents, thereby underscoring its robustness and specificity. In another study, a nanohybrid-based immunosensor was developed to detect the *H. pylori* *Baba* antigen. This biosensor utilized an electrode loaded with palladium nanoparticles (Pd) through electrodeposition to create rGO and poly(3,4-ethylenedioxythiophene) (PEDOT). This nanohybrid platform enhanced the surface area and electroconductivity for the immobilization of antibodies and the *Baba* antigen.⁸⁵ The immunosensor demonstrated the capability to detect *H. pylori* antigens in a linear range of $0.2\text{--}20 \text{ ng mL}^{-1}$ with an LOD of 0.2 ng mL^{-1} . Computational simulations confirmed stable antigen–antibody interactions, emphasizing the sensor's functionality in the electrochemical detection of *H. pylori*. This immunosensor exhibited high specificity, sensitivity, and reproducibility,

making it a promising tool for accurate detection of the *H. pylori* *Baba* antigen.

Furthermore, a label-free electrochemical immunosensor was fabricated for rapid detection of *H. pylori*. This sensor was based on the combination of titanium oxide, carboxylated multi-walled carbon nanotubes, and polyindole carboxylic acid composites.¹⁴⁷ The sensor demonstrated an LOD of 0.1 ng mL^{-1} in a dynamic linear range of $0.1\text{--}8.0 \text{ ng mL}^{-1}$, having high sensitivity and selectivity. The study involved testing for *H. pylori* detection in five human stool specimens, showing good results with suitable accuracy.^{147,148} The analysis of these studies highlights the potential of electrochemical biosensors for rapid, accurate, and timely detection of *H. pylori* infection. The rapid and accurate detection of *H. pylori* infection facilitated by these electrochemical biosensors has led to timely therapeutic interventions, ultimately improving patient outcomes and reducing the risk of associated complications. In another study, a titanium carbide ($\text{Ti}_3\text{C}_2\text{T}_x$) MXene-based electrochemical biosensor was developed for the detection of antibodies in serum samples. The biosensor exhibited higher sensitivity with an LOD of 0.1 ng mL^{-1} and better selectivity compared to conventional serological tests like ELISA.⁷⁹ The superior performance was attributed to the exceptional conductivity, high bandgap, large surface area, and tunable surface chemistry of $\text{Ti}_3\text{C}_2\text{T}_x$ MXene, enabling efficient antibody immobilization and signal transduction. Furthermore, an immunosensor was developed for CagA antigen detection.



The gold electrode was subsequently modified by platinum nanoparticles, poly(3,4-ethylenedioxythiophene) (PEDOT), and reduced graphene oxide (rGO) for immobilization of the CagA antigen on this electrode. The fabricated device was used for immunosensing through interaction between antigen and antibodies present in serum samples. The linear range of the biosensor was calculated as 0.1 ng mL⁻¹ to 30 ng mL⁻¹ and LOD as 0.1 ng mL⁻¹.⁸⁰

Similar to immuno-based biosensors, researchers have developed nucleic acid based sensors, such as DNA or aptamer based biosensors, to detect *H. pylori*. For example, the DNA-based sensors rely on the hybridization of the single-stranded DNA probe with its complementary DNA target sequence. These DNA-based biosensing platforms are advantageous due to their high chemical stability and ease of modification, and have attracted the focus of many researchers developing biosensors aimed at detecting *H. pylori*.¹⁴⁹ The ability to tailor the DNA probes and transduce hybridization events into measurable outputs has led to widespread exploration of DNA biosensors for detection of *H. pylori*. Hajihosseini *et al.* developed a sensitive electrochemical DNA biosensor for detecting *H. pylori* utilizing a single-stranded DNA probe immobilized on a graphene oxide/gold (GO/Au) nanocomposite modified glassy carbon electrode. The biosensor demonstrated a linear range of 60 pM to 600 pM with an LOD of 27 pM to detect *H. pylori*.¹⁵⁰

del Pozo *et al.* employed a surface-based approach to study DNA-based electrochemical interactions with redox-active osmium complexes containing 1,10-phenanthroline-5,6-dione ligands. In this approach, gold electrodes were functionalized with DNA probes and osmium reporter molecules like [Os(bpy)2(phen-dione)]^{3+/2+} enabling accumulation within the DNA layer. Voltammetric charge quantified the bound osmium complexes, exploiting the phen-dione quinone's low-potential redox activity for electrochemical DNA sensing. A single-stranded *H. pylori* DNA probe was linked with thiol and immobilized on gold, and hybridization with its complement allowed osmium accumulation within the dsDNA. The platform enabled quantification of *H. pylori* sequences ranging from 5 to 20 pM, with a detection limit of 6 pM.¹⁵¹ In another study, Asadzadeh-Firouzabadi *et al.* demonstrated a novel electrochemical DNA hybridization biosensor with a signal range from 20.0 to 410.0 nM for *H. pylori* complementary DNA, achieving a detection limit of 7.2 nM. This sensing platform utilized chlorogenic acid (CGA) as the electroactive indicator. The detection mechanism of the sensor relied on interaction of CGA with an immobilized 18-mer *H. pylori* alkanethiol DNA probe and its hybridized form on a self-assembled gold electrode.⁸⁴

In addition to DNA, researchers employ small oligonucleotides known as aptamers for creating nucleic acid-based electrochemical biosensors. These aptamers exhibit exceptional binding affinity for their targets; they are selected from randomized libraries and are capable of recognizing and binding to particular protein regions with high specificity. For example, Yadav and colleagues developed an aptasensor by utilizing graphitic carbon nitride (g-C₃N₄) with 3-(aminopropyl)triethoxysilane (APTES). *H. pylori* CagA-specific aptamers were covalently

conjugated to the modified electrode enabling the detection of CagA with 1.98 μA ng mL⁻¹ cm⁻¹ sensitivity and 0.1–160 ng mL⁻¹ linear range. The aptasensor demonstrated an LOD of 0.017 ng mL⁻¹. In this study, the aptasensor developed exhibited exceptional specificity toward the *H. pylori* CagA toxin, enabling highly stable gastric cancer detection for point-of-care applications.¹⁵²

3.3.2. Optical-based biosensors. Recent advancements in optical biosensors have significantly enhanced the detection of infections by offering high sensitivity, rapid response times, and multiplexed detection capabilities. One notable advancement is the development of plasmonic biosensors, which leverage the unique optical properties of plasmonic nanostructures to detect biomolecular interactions with exquisite sensitivity.¹⁴⁷ For example, surface plasmon resonance (SPR) biosensors utilize the evanescent field generated by the interaction of light with a metal-dielectric interface to monitor changes in the refractive index upon biomolecular binding.¹⁴⁷ Recent studies have demonstrated the application of SPR biosensors for the detection of various infectious agents, including bacteria, viruses, and parasites, by immobilizing specific capture molecules, such as antibodies or aptamers, onto the sensor surface (Fig. 4). These plasmonic biosensors offer label-free detection, real-time monitoring, and high-throughput capabilities, making them invaluable tools for infectious disease diagnosis and surveillance.

In addition to plasmonic biosensors, advancements in photonic and photonic crystal biosensors have enabled highly sensitive and selective detection of infections. Photonic biosensors exploit the principles of light propagation and modulation within photonic structures, such as waveguides, resonators, and gratings, to detect changes in the refractive index or optical properties induced by biomolecular interactions. For instance, photonic crystal biosensors utilize periodic nanostructures with tailored optical properties to achieve enhanced sensitivity and specificity in detecting infectious agents.¹⁴⁷ Studies have demonstrated the application of photonic crystal biosensors for the label-free detection of bacteria, viruses, and toxins in clinical samples, offering advantages such as rapid analysis, low sample consumption, and compatibility with microfluidic systems. These photonic biosensors hold promise for point-of-care diagnostics, environmental monitoring, and biosecurity applications, facilitating early detection and containment of infectious diseases.

Furthermore, the integration of advanced materials, such as graphene, quantum dots, and metal nanoparticles, into optical biosensors has led to significant improvements in sensitivity, signal amplification, and multiplexed detection capabilities. Graphene-based biosensors, for example, offer unique optical and electrical properties that enable highly sensitive detection of biomolecular interactions.¹⁵³ Similarly, quantum dot-based biosensors utilize semiconductor nanoparticles with tunable optical properties to achieve ultrasensitive detection of biomolecules. In a study, a FRET (fluorescence resonance energy transfer) based biosensor was developed for sensitive determination of *H. pylori*. Two types of oligonucleotide probes were labeled with cadmium telluride (CdTe) QDs (the donor) and





Fig. 4 Advanced design of optical label-free biosensors for detection of *H. pylori* biomarkers in biological fluids.

5-carboxytetramethylrhodamine (the acceptor) respectively. Following the donor's photoexcitation, the donor and acceptor's proximity is used to measure FRET. The probes were not able to bind in the absence of target analytes and, subsequently, there was no emission. The fabricated biosensor was simple, rapid, and highly efficient for the detection of homogeneous DNA.¹⁵⁴ Israa *et al.* developed an optical biosensor based on a multi-mode no core fiber to determine the minimum inhibitory concentrations (MICs) of antibiotics against *H. pylori*. In this study, the biosensor used an inline Mach-Zehnder interferometer to measure the bacteria in the BHI broth turbidity stabilizer and compared the results with traditional methods.¹²⁵ The developed optical biosensor was presented as a facile, rapid, and effective method for bacterial detection in comparison to the conventional methods.

Shahrashoob *et al.* reported a novel optics-based biosensor to detect a *H. pylori* urease-labeled probe. The probe was immobilized on APTES-activated glass-bound gold nanoparticles. The hybridization between the probe and the target analyte resulted in a decrease in the optical density corresponding to the analyte concentration. This optics-based approach achieved a remarkable limit of detection of 0.5 nM for *H. pylori*.¹⁵⁵ In a recent study, Rong *et al.* proposed a novel optical biosensor utilizing porous silicon (PSi) nanomaterial and photonic Tamm plasmon polariton (TPP) surface states to detect *H. pylori CagA*. The findings of this study demonstrated the biosensor's high sensitivity of 100 pm (ng mL⁻¹)⁻¹, low detection limit of 0.01 ng mL⁻¹, and excellent specificity with a positive-to-negative ratio exceeding six. These promising performance metrics suggest the potential of the TPP biosensor for reliable diagnosis of *H. pylori* infection in both laboratory and point-of-care testing settings.¹⁵⁶

3.3.3. Colorimetric-based biosensors. 2D nanoparticle-based colorimetric biosensors are valuable tools in the detection of diseases due to their simplicity, rapidity, and cost-effectiveness. These biosensors rely on the principle of color change to indicate the presence of target biomolecules associated with the disease of interest. When the target biomolecules bind to the immobilized bioreceptors on the nanoparticle surface, it induces aggregation or dispersion of the nanoparticles, leading to a change in their optical properties, such as color or absorbance. This change can be easily visualized with the naked eye or measured using simple spectrophotometric techniques, enabling qualitative or quantitative detection of diseases.¹⁵⁷ Fig. 5 shows the main elements of a colorimetric biosensor and the steps in the diagnosis of *H. pylori*. Recent advancements in the field of 2D nanoparticle-based colorimetric biosensors have significantly improved their performance and applicability for the detection of *H. pylori* infection. Recently, a colorimetric biosensor based on the peroxidase-like activity of 2D nanomaterials which detects antibodies in serum has been developed for infectious disease detection.¹⁵⁷ The Co₃O₄ nanosheets used exhibited intrinsic peroxidase-like activity, which catalyzed the oxidation of a colorless substrate (3,3',5,5'-tetramethylbenzidine) in the presence of hydrogen peroxide, resulting in a blue-colored product.¹⁵⁸ The presence of specific antibodies in the sample induced a color change due to the inhibition of the peroxidase-like activity of Co₃O₄. This biosensor exhibited high sensitivity, with a detection limit of 0.5 ng mL⁻¹, and excellent selectivity against other interfering antibodies.¹⁵⁷

Recently, Suaifan *et al.* conducted a study focusing on the creation of a colorimetric nanomaterial-based paper biosensor specifically designed for the detection of *H. pylori* using





Fig. 5 Diagram of main elements of a colorimetric biosensor and the steps in the diagnosis of *H. pylori*. The binding of the target to the immobilized bioreceptors on the nanoparticle surface induces aggregation or dispersion of the nanoparticles which results in a change in color.

extracellular proteases as biomarkers.¹⁵⁹ This innovative biosensor employs a unique substrate that is functionalized with magnetic nanobeads and integrated into a gold sensing platform. The biosensor demonstrated a limit of detection (LOD) of 100 CFU mL⁻¹,¹⁵⁹ while also exhibiting notable selectivity and stability. Furthermore, its efficacy in detecting *H. pylori* in clinical specimens was evaluated, yielding promising results regarding both sensitivity and specificity.

Fei *et al.* developed a colorimetric biosensing assay to detect *H. pylori* in stool samples. The biosensing assay utilized oligonucleotide probes impregnated with gold nanoparticles together with aptamers specific for *H. pylori*.¹⁶⁰ The aggregation and deaggregation helped to achieve the detection process with a detection limit of 25 CFU mL⁻¹. This assay offers a promising method for clinical field detection of *H. pylori*, as it is noninvasive and can be used for both qualitative and quantitative detection. As research in this field continues to progress, we can expect even more sensitive, selective, and user-friendly colorimetric biosensors for the rapid and accurate detection of *H. pylori* and other infectious diseases, ultimately contributing to improved patient outcomes and public health management.

3.3.4. Field-effect transistor biosensors. Field effect transistor (FET) biosensors have been developed for the detection of *H. pylori*. These biosensors utilize FET structures, such as chemical FETs and ion-sensitive FETs, to detect the presence of *H. pylori* by measuring changes in electrical current or pH levels. The signal transduction interface between the electrode surface and the biological sample is crucial in FET biosensors because it captures target ions or biomolecules and converts

their electrical characteristics and electrochemical reactions into output signals.¹⁶¹ The FET biosensors can be used in various forms, including endoscopic tube-tip sensors and solid-phase capillary-tube sensors. The endoscopic tube-tip sensor allows for on-site detection of *H. pylori* by measuring urease activity on the gastric wall using a pH-FET.¹⁶¹ The solid-phase capillary-tube sensor selectively captures *H. pylori*'s urease in gastric mucus and measures its activity using a pH-FET. These FET biosensors have shown promising results in terms of sensitivity and specificity, with clinical evaluations reporting sensitivity ranging from 89% to 100% and specificity ranging from 86% to 98%.^{161,162} Recently, a FET biosensor for *H. pylori* detection based on a chemical field-effect transistor was developed. The sensitivity of the developed device for detecting *H. pylori* strains ranged from 2.1 kHz ppm⁻¹ to 3.4 kHz ppm⁻¹.¹²² The developed frequency transducer shows promise for detecting *H. pylori* strains, with its high sensitivity and ability to accurately determine gas concentration changes.

Researchers have also explored the integration of 2D nanoparticle-based FET biosensors with microfluidic and lab-on-chip devices to enhance their performance and facilitate point-of-care (POC) diagnostics. A microfluidic chip-based FET biosensor has been developed to quantify the presence of pathogenic biomarkers.¹⁶³ These biosensors have been shown to utilize complementary metal oxide or carbon semiconductor as the channel material in the FET device, and specific biomarkers such as *H. pylori CagA* or *Baba* can be immobilized on its surface. The binding of biomarkers will induce a change in the electrical properties of the metal oxide or carbon channel,



resulting in a measurable change in the source–drain current (Fig. 6). These integrated biosensors exhibit high sensitivity, with a lower detection limit and they demonstrate excellent performance in the analysis of clinical samples.^{163,164} The advancements in 2D nanoparticle-based FET biosensors offer several advantages, including label-free detection, high sensitivity, and the potential for real-time monitoring. As research in this field continues to progress, we can expect even more sensitive, selective, and user-friendly FET biosensors for the rapid and accurate detection of *H. pylori* and other infectious diseases, ultimately contributing to improved patient outcomes and public health management.

3.3.5. Piezoelectric biosensors. Piezoelectric biosensors have been proposed for the detection of *H. pylori* infection. These biosensors involve the use of piezoelectric materials such as zinc oxide and polyvinylidene fluoride (PVDF) as well as electrode materials like aluminium, gold, and graphite.¹²⁶ The biosensor designs are coated with a uniform layer of antibodies, and simulations using COMSOL Multiphysics software are performed to measure the voltages generated in the presence and absence of an antigen.¹⁶⁵ When the target biomolecules bind to the immobilized receptors on the nanoparticle surface, they induce a mechanical strain or mass change in the piezoelectric material, resulting in a change in its electrical properties, such as frequency or impedance. This change can be quantified using piezoelectric transducers, such as quartz crystal microbalances (QCMs) or surface acoustic wave (SAW) devices, enabling sensitive and real-time detection of diseases. The difference in voltages produced can be used to detect the presence of *H. pylori*. This approach offers a rapid and noninvasive method

for the detection of *H. pylori*, which is important for ensuring the health and safety of individuals.^{126,165}

In a reported study, an enzymatically amplified piezoelectric immunosensor was developed for the sensitive detection of antibodies against the *H. pylori* pathogen. In this sensor, a sandwich assay format on AT-cut quartz crystal resonators operating at 10 MHz frequency was employed. Recombinant *H. pylori* antigen proteins were then immobilized onto the quartz crystal surface serving as capture probes for the corresponding antibodies present in patient serum samples. It was reported that the study employed anti-human IgG secondary antibody probes specifically binding to the human antibodies previously captured on the antigen-modified surface, forming sandwiched immunocomplexes. This secondary antibody binding event significantly improved the positive-to-negative signal ratio, which was increased to 4.1–5.0. Moreover, the horseradish peroxidase enzyme was conjugated with anti-IgG probes to achieve an additional signal amplification. This enzymatic amplification strategy resulted in a striking enhancement of sensitivity, enabling reliable positive detection even in serum samples diluted up to 1/100. Through this multi-pronged approach, the piezoelectric immunosensor overcomes the obstacles posed by low antibody levels in clinical specimens, paving the way for efficient serological diagnosis of *H. pylori* infections.¹²³ Fig. 7 shows a piezoelectric immunosensor for the determination of *H. pylori* and monoclonal Abs specific for *H. pylori*.

3.4. Design considerations for optimizing biosensor performance

The primary design goal for biosensors intended for the clinical detection of *H. pylori* infections is to achieve high sensitivity, specificity, selectivity, and reliability. Several interdependent



Fig. 6 Carbon nanowire-based field-effect transistor biosensor for the detection of *H. pylori*. The binding of the analyte with biorecognition elements will induce a change in the electrical properties of the carbon nanowire channel, resulting in a measurable change in the source–drain current.



factors influence biosensor's analytical performance and require careful optimization during the design phase. One critical aspect is the selection and immobilization of biorecognition elements onto the biosensor surface.^{64,66} Optimizing the design of biosensors for *H. pylori* detection plays a pivotal role in translating scientific innovations into effective clinical outcomes. Emerging research underscores the crucial role of choosing suitable transducers and bio-recognition elements, as these components directly impact the sensitivity and specificity of the biosensors. For example, surface plasmon resonance (SPR) biosensors have exhibited exceptional sensitivity, capable of detecting *H. pylori* at concentrations as low as 200 CFU mL⁻¹, being significantly more sensitive compared to conventional diagnostic approaches like the rapid urease test (RUT) that require much higher bacterial concentrations to achieve reliable detection.^{166,167} For *H. pylori* detection, antibodies, aptamers, and DNA probes specific to *H. pylori* antigens or DNA sequences are commonly used as biorecognition elements. Antibodies, though highly specific, can suffer from batch-to-batch variability, instability and high costs. Aptamers are increasingly preferred as robust, reusable synthetic receptors but require stringent selection strategies.^{70,168} DNA/PNA probes enable excellent specificity but are unsuitable for detecting proteins/whole cells.¹⁶⁹ For instance, a recent study reported the development of an ultrasensitive electrochemical aptasensor for *H. pylori* detection. This sensor is based on surface modified graphene oxide doped gold nanoparticles conjugated polythiophene. The linear behavior of the modified electrode against a wide range of concentrations of *H. Pylori*, shows the reliability of the developed biosensor, with a limit of detection (LOD) of 0.0080 μM for impedance and 0.0067 μM for square wave voltammetry curves.¹¹⁰ The optimized immobilization strategy of the aptamer on the polythiophene (PTP)-modified electrode resulted in enhanced sensitivity and stability of the biosensor,

enabling the rapid and reliable detection of *H. pylori* antigens in clinical samples.¹¹⁰ Surface chemistries utilized for bioconjugation also impact orientation, steric hindrance and bioactivity.

Furthermore, the choice of the transducer substrate influences biosensor sensitivity. Electrochemical transducers, such as screen-printed electrodes (SPEs) and field-effect transistors (FETs), are commonly employed for *H. pylori* detection due to their high sensitivity, low cost, and compatibility with miniaturized devices for point-of-care testing. Recent advancements in signal amplification techniques, such as enzymatic amplification, nanoparticle-based amplification, and signal enhancement strategies, have further improved the detection limits and dynamic range of biosensors for *H. pylori* detection. For example, Chen *et al.* developed a MoS₂-based biosensor for exosome detection, where signal amplification was achieved through the use of horseradish peroxidase (HRP)-exosome probes and the tetramethylbenzidine (TMB) substrate, resulting in significantly enhanced sensitivity and specificity compared to conventional methods.⁹³

In addition, minimizing non-specific binding and matrix interference from complex clinical samples is another critical aspect in biosensor design. Immobilizing biorecognition elements in oriented configurations helps reduce steric hindrances. Incorporating blocking agents like BSA, ethanolamine, and pyrene layers prevents non-specific adsorption. Wang *et al.* developed a 3D reduced graphene oxide-polypyrrole immunosensor using aryldiazonium salt grafts to create antifouling surfaces, enhancing specificity against interfering proteins.⁹² 2D nanomaterials can initially be more expensive than conventional materials. However, these costs can decrease as mass production and miniaturization become more feasible, leading to potential economies of scale.¹¹⁹ In addition, maintenance costs for nanomaterial-based biosensors tend to be lower than for traditional equipment, as they are typically more robust,



Fig. 7 Schematic diagram showing the mechanism of the piezoelectric immunosensor for the detection of *H. pylori* from human serum samples.



portable, and easier to use, particularly in resource-limited settings. Surface functionalization with zwitterionic polymers like poly(carboxybetaine) creates highly antifouling yet biocompatible biosensor interfaces.⁹⁰ Other antifouling materials include polyethylene glycol, dextran, and polyoxazoline.⁹⁰ The use of these materials increases the upfront costs associated with developing the biosensors. Advanced designs employing microfluidic, paper-based, and lab-on-chip devices will minimize sample volumes, improve reaction kinetics and enable instrument-free operation. A recent 3D-printed microfluidic electrochemical cytosensor detected cancer cells directly from gastric biopsy specimens without complex labeling.⁹¹ Similarly, a low-cost paper-based vertical flow assay incorporated antibody-conjugated carbon nanotubes for visual immunoassay.¹³² Smart design strategies like these that synergistically optimize biorecognition elements, transducers, surface chemistry and device architecture are key to realizing high-performance biosensors. These designs can be integrated for robust *H. pylori* diagnosis at the point-of-care.

3.5. Industrial and commercialization viability of 2D nanomaterial-based biosensors

The industrialization potential and commercial viability of 2D nanomaterial-based biosensors for *H. pylori* detection are significant, especially given the growing demand for rapid, sensitive, and cost-efficient diagnostic tools in healthcare. The use of 2D nanomaterials in biosensors can dramatically lower the detection limits for *H. pylori*, which is essential for early diagnosis and treatment of gastrointestinal disorders.^{125,151} A nanohybrid-based immunosensor was developed for the detection of the *H. pylori* BabA antigen.⁸⁵ The reduced graphene oxide/poly(3,4-ethylenedioxythiophene) immunosensor developed demonstrated a strong ability to detect *H. pylori* antigens over a linear detection range of 0.2 to 20 ng mL⁻¹, with a low limit of detection of 0.2 ng mL⁻¹.⁸⁵ This sensitivity highlights its potential for industrial and commercialization purposes enhancing accurate and prompt diagnosis of *H. pylori*-related infections. Additionally, the simplicity and affordability of electrochemical biosensors made with 2D materials make them attractive alternatives to conventional diagnostic methods, which are often complicated and require specialized skills.^{20,120,170} Zhang *et al.* introduced a novel glucose sensor utilizing enzyme-modified whole-graphene solution-gated transistors (SGGTs).¹⁷¹ The sensor operates through the detection of hydrogen peroxide (H₂O₂) produced during glucose oxidation, which occurs near the gate. This innovative biosensor achieves detection limits below 0.5 μM, demonstrating excellent selectivity for non-invasive glucose monitoring in bodily fluids.¹⁷¹ The transconductance of these devices is significantly higher – by two orders of magnitude – compared to that of traditional silicon field-effect transistors, which is a critical factor for their heightened sensitivity.^{95,162,171} Additionally, the fabrication process is straightforward and cost-effective, enhancing commercialization potential for practical application in biosensing technologies.

Point-of-care testing is gaining traction in healthcare settings, particularly in under-resourced areas, due to its portable and user-friendly nature, which allows for immediate results. This trend is bolstered by investments from governments and

health organizations aimed at advancing disease detection and management technologies, thereby expanding the biosensor market. The biosensor sector is projected to grow significantly, with estimates suggesting a market size of USD 24.12 billion in 2024, potentially reaching USD 36.55 billion by 2029, and reflecting a compound annual growth rate (CAGR) of 8.67% during this period.¹⁷² The rise in chronic diseases, including peptic ulcers, diabetes, and cancer, underscores the necessity for continuous monitoring of *H. pylori* infection, blood glucose levels, and cancer cell proliferation.^{93,157,173} This demand drives the development of implantable and wearable biosensors capable of providing accurate real-time monitoring. Specifically, the need for reliable biosensors for *H. pylori* detection is critical, as early diagnosis and ongoing management of gastritis are vital for preventing severe complications associated with the infection.

The widespread commercialization and adoption of 2D nanomaterial-based biosensors face challenges, including the lack of standardized safety and efficacy guidelines for nanomaterials, which complicates market entry due to the demands of the regulatory bodies for comprehensive risk assessments.^{114,121} Although 2D nanomaterials enhance sensor performance, their synthesis and integration can be costly, limiting accessibility, particularly in resource-limited settings. To ensure practical applications, issues such as the long-term stability, biocompatibility, and consistent performance of nanomaterials must be addressed. Furthermore, the competitive biosensor market requires new entrants to demonstrate clear advantages over existing technologies to gain traction and secure a foothold in the industry.

4. Challenges and future perspectives

4.1. Challenges in fabrication, integration and biocompatibility

The integration of 2D nanoparticles in biosensors for the detection of *H. pylori* and other infectious diseases has shown promising results, but it is not without challenges. Different studies have shown that researchers encounter notable challenges during the fabrication and integration processes. The controlled synthesis and functionalization of 2D nanoparticles is one of the significant challenges that nanosensing technology is facing. These nanomaterials often exhibit diverse physical and chemical properties which make it difficult to achieve consistent and reproducible characteristics. For instance, the fabrication of GO and MoS₂ nanosheets in biosensors displayed varying degrees of oxidation, defect density, and surface functional groups, which impacted their interaction with biorecognition elements and target analytes.^{174,175} This means that multi-step oxidation processes are required to improve the reproducibility and performance of the biosensor with such nanomaterials. Studies have shown that the integration of 2D nanoparticles into some biosensing platforms often requires complex fabrication processes and specialized equipment. Savannah *et al.* investigated the development of a graphene-field effect biosensor to enable real-time, quantitative



detection of the Zika virus.¹⁷⁶ The use of graphene nanosheets offered advantages such as high surface area and excellent biocompatibility, but challenges arose concerning the reproducibility of the fabrication process and the optimization of graphene nanosheet dispersion in the biosensor.

Another challenge lies in the effective immobilization and orientation of biorecognition elements on the surface of 2D nanoparticles. Improper immobilization can lead to reduced binding affinity, steric hindrance, and decreased sensitivity of the biosensor. The integration of MoS₂ nanosheets with a specific DNA aptamer encountered challenges in achieving optimal orientation and spacing of the aptamer on the MoS₂ surface, which initially resulted in reduced target binding and signal transduction.^{174,175} In addition to fabrication challenges, the integration of 2D nanoparticles in biosensors for infectious disease detection can be a challenge due to biocompatibility, stability, and potential toxicity properties of the 2D nanomaterial in question.^{100,102,104,114,130} Nanomaterials may interact with biological systems in unexpected ways, potentially causing adverse effects or interfering with the performance of the biosensor. Extensive testing and optimization are required to ensure the biosafety and long-term stability of these biosensor systems, especially for *in vivo* applications.

In addition, biocompatibility presents a significant challenge in the development of 2D nanomaterial-based biosensors, as these materials can elicit immune responses or exhibit cytotoxic effects when in contact with biological systems.^{109,177} To address these concerns, one effective approach is to functionalize nanomaterials with biocompatible polymers or biomolecules, enhancing their compatibility with biological components while reducing toxicity. Over the past years, polymer-based biosensors have demonstrated success in detecting various analytes, including glucose, proteases, and hydrogen peroxide.¹⁷⁸ Recent research has highlighted their application in the detection of *H. pylori*. For instance, Saber Mirzaei *et al.* developed an ultrasensitive electrochemical aptasensor utilizing surface-modified graphene oxide combined with gold nanoparticles and conjugated polythiophene polymer for *H. pylori* detection.¹¹⁰ This innovative biosensor exhibited impressive limits of detection, reaching 0.0080 μM for impedance measurements and 0.0067 μM for square wave voltammetry (SWV). Surface modifications with biocompatible polymers can protect nanomaterials from direct exposure, further enhancing their safety in medical applications.

4.2. Challenges related to stability, reproducibility and sensitivity

The stability and reproducibility of 2D nanoparticles in biosensing platforms still remain a challenge. 2D nanomaterials often exhibit high surface energy and reactivity, which makes them susceptible to degradation or alteration over time. Investigations on the stability of graphene oxide nanosheets used in an electrochemical biosensor found that the nanosheets underwent structural changes and oxidation over time, leading to a gradual decrease in the sensitivity and accuracy of the biosensor.^{114,130,174} In this study, the use of protective coatings and encapsulation techniques enhanced the long-term stability

of the GO nanosheets, but these approaches added complexity and cost to the fabrication process. Challenges with the reproducibility of the physicochemical properties of 2D nanoparticles can significantly impact the performance and reliability of biosensors. The reproducibility of biosensor fabrication using 2D nanoparticles is also hindered by variations in material properties, synthesis methods, and environmental factors.^{126,129,130,153,160} This significantly affects the acceptability and commercialization of biosensing technologies during epidemics. Additionally, the dispersion of 2D nanoparticles within sensing matrices poses a significant challenge, as agglomeration can compromise the performance of the biosensor and result in unreliable detection outcomes.¹⁷⁹ The variations in the size, thickness, and defect density of the MoS₂ nanosheets led to inconsistencies in the sensitivity and selectivity of the biosensors developed using such nanosheets. Exposure of 2D nanoparticle-based biosensors to harsh environmental conditions such as temperature fluctuations, humidity, and potential contaminants can affect their performance and reliability and result in loss of sensitivity over prolonged use. Some 2D nanoparticles exhibit high magnetic properties which poses high chances of contamination with other magnetic particles during exposure to the environment. Several studies have highlighted that the instability of graphene-based nanomaterials in physiological environments can lead to degradation and loss of sensing functionality over time.^{100,180-183}

Wu *et al.* investigated the stability of a surface-enhanced Raman scattering (SERS) biosensor based on silver nanoparticles and found that the SERS biosensor exhibited a gradual decrease in sensitivity and signal intensity when exposed to high temperatures or humid conditions, potentially limiting its practical applications in resource-limited settings or field-based diagnostics.¹⁸⁴ Furthermore, different studies have noted that unreliable or inconsistent performance due to stability, scalability, and reproducibility related issues can lead to inaccurate diagnoses, potentially compromising public health efforts and undermining the confidence of healthcare professionals and regulatory authorities in these biosensing platforms. The challenges together are multifaceted and require concerted efforts from multidisciplinary teams to unlock the full potential of these advanced biosensing platforms, enabling rapid and accurate disease detection during epidemics and enhancing public health preparedness on a global scale.

4.3. Strategies for improving biosensor performance

Recent advancements in biosensors for the detection of infectious diseases, particularly targeting *H. pylori*, have seen a surge in interest due to the pressing need for rapid and accurate diagnostic tools. The optimization of biosensor performance for the detection of *H. pylori* is essential for enhancing diagnostic accuracy and clinical applicability. Recent studies have highlighted several strategies that can significantly improve biosensor functionality, bridging scientific advancements with real-world clinical implications.^{87,166} Researchers have been actively exploring strategies to enhance the performance of biosensors, including surface functionalization, bioconjugation, hybrid nanostructures, and nanocomposites. These approaches



aim to improve sensitivity, selectivity, and integration with microfluidic and lab-on-chip systems, ultimately enabling more efficient and reliable *H. pylori* detection. Surface functionalization and bioconjugation have emerged as powerful strategies to improve biosensor performance. A number of studies have demonstrated the promising potential of graphene oxide nanosheet-based electrochemical biosensors in detecting specific bacterial strains with high sensitivity and selectivity. Hajihosseini *et al.* developed biosensors utilizing GO-based platforms that could sensitively and selectively detect *H. pylori*. These biosensors demonstrated exceptional performance, with detection limits as low as 27.0 pM for *H. pylori*.¹⁵⁰ Additionally, Duan *et al.* demonstrated the potential of a GO-based aptasensor for detecting *Salmonella typhimurium*, with a detection limit of 100 CFU mL⁻¹.¹⁸⁵ This aptamer-functionalized GO biosensor exhibited remarkable stability and reusability, making it a promising platform for point-of-care diagnostics.

Another strategy involves the development of hybrid nanostructures and nanocomposites, which combine the unique properties of different nanomaterials to create synergistic effects. Saxena *et al.* reported the fabrication of an electrochemical immunosensor, which detected the *VacA* gene, on a gold electrode decorated with graphitic carbon nitride/zinc oxide.¹⁸⁵ The g-C₃N₄/ZnO nanocomposite was functionalized with specific antibodies, enabling selective capture and detection of anti-*H. pylori VacA* antigens in patient serum samples. The combination of graphitic carbon nitride with zinc oxide enhanced the performance of the biosensor which showed a sensitivity of 0.3 μA ng⁻¹ mL⁻¹ in a linear detection range of 0.1 to 12.8 ng mL⁻¹.^{86,117} Integrating biosensors with microfluidic and lab-on-chip systems has also emerged as a promising strategy for enhancing biosensor's ability to detect *H. pylori* and other infectious agents. These integrated systems offer several advantages, including reduced sample and reagent volumes, automated sample handling, and the potential for multiplexed analysis. Yao Xie *et al.* demonstrated the development of a microfluidic electrochemical biosensor for the detection of gastric cancer biomarkers. In this study, an electrochemical microfluidic chip that combines multiple biomarkers, including carbohydrate antigen 19-9 (CA19-9), *H. pylori CagA* protein (H.P.), P53 oncoprotein (P53), pepsinogen I (PG I), and PG-II, for the early detection of gastric cancer was developed.¹⁸⁶ The developed sensor offered better sensitivity compared to enzyme-linked immunosorbent assay results.

In addition, the development of robust encapsulation or surface modification techniques to improve the stability and durability of 2D nanoparticles in biosensors is a promising strategy to improve biosensor performance. Studies have demonstrated the use of a polymer-based encapsulation method to enhance the long-term stability and reproducibility of GO nanosheets in an electrochemical biosensor for the detection of *H. pylori* antibodies.^{117,150,187} The encapsulated GO nanosheets exhibited minimal degradation and consistent performance over an extended period, offering a potential solution for commercialization and widespread adoption. Furthermore, the development of standardized protocols and quality control measures for the synthesis, characterization,

and integration of 2D nanoparticles in biosensors is crucial for ensuring reproducibility and consistent performance. Collaborative efforts among research institutions, industry partners, and regulatory bodies are essential to establish guidelines and best practices for the fabrication and validation of these biosensing platforms, facilitating their acceptability and commercialization during epidemics.

4.4. Future directions and conclusions

The future of 2D nanoparticle-based biosensors for the detection of *H. pylori* holds immense potential, driven by advancements in multiplexed detection, point-of-care (POC) applications, integration with wearable and implantable devices, CRISPR-Cas technology, and the incorporation of machine learning (ML) and artificial intelligence (AI) algorithms. One significant direction for future research involves the development of multiplexed biosensors capable of simultaneously detecting multiple biomarkers associated with *H. pylori* infection. By incorporating a variety of recognition elements, such as antibodies, aptamers, and DNA probes, into the biosensor platform, researchers can enhance the specificity and accuracy of detection while enabling comprehensive profiling of the pathogen's biomolecular signatures. For example, recent studies have demonstrated the feasibility of multiplexed detection using arrays of functionalized graphene-based nanomaterials, enabling the simultaneous detection of *H. pylori* antigens, DNA sequences, and metabolic biomarkers.^{188–190} This multiplexed approach not only streamlines diagnostic workflows but also facilitates personalized treatment strategies tailored to individual patient profiles, thereby improving clinical outcomes and reducing healthcare costs associated with *H. pylori*-related diseases.

Furthermore, the integration of 2D nanoparticle-based biosensors with POC devices represents a paradigm shift in biosensing technology, enabling rapid and decentralized diagnosis of *H. pylori* infection at the point of need. POC biosensors leverage miniaturized and portable platforms, such as microfluidic chips and smartphone-based devices, to enable real-time analysis of clinical samples in resource-limited settings. For instance, recent advancements in microfluidic-based biosensors have enabled the development of compact and user-friendly POC devices for the detection of pathogenic bacterial cells, antigens and DNA in saliva, blood, and stool samples.^{186,191,192} These POC biosensors offer significant advantages in terms of accessibility, affordability, and scalability, making them invaluable tools for community-based screening, outbreak surveillance, and remote monitoring of *H. pylori* infection. In addition, wearable and implantable devices integrated with biosensors hold promise for continuous and non-invasive monitoring of *H. pylori* infection dynamics in real-time. Wearable biosensors, such as smart patches and wearable bands, can detect biomarkers associated with *H. pylori* infection from bodily fluids, sweat, or breath, providing valuable insights into disease progression and treatment response.¹⁹³ Implantable biosensors, on the other hand, offer long-term monitoring capabilities by interfacing directly with the host's physiological environment, enabling early detection of *H. pylori*-related complications, such as gastric ulcers and gastric cancer.^{194,195} These integrated diagnostic strategies pave the way for personalized and



preemptive healthcare interventions, ultimately improving patient outcomes and reducing the burden of *H. pylori*-related diseases on healthcare systems worldwide.

CRISPR-Cas technology also offers promising results for biosensor-based disease diagnosis. This technology involves precision in targeting specific DNA or RNA sequences which enables highly sensitive and specific detection of pathogens or genetic mutations associated with diseases.¹⁹⁶ A CRISPR/Cas12a-assisted array has been developed for the analysis of *H. pylori* DNA in saliva.¹⁹⁷ This method utilizes the CRISPR/Cas12a system to detect specific genes of *H. pylori*, such as the 16S rDNA, *CagA*, and *VacA* genes.^{22,197,198} The assay process can be performed at a single temperature in about 30 minutes, making it rapid and convenient. The sensitivity of the method is 2 copies per μL and the LOD is about 50 copies.¹⁹⁹ The CRISPR/Cas12a system has been shown to have high specificity for target recognition. The detection results can be displayed using lateral flow strips, allowing for easy interpretation.^{22,196–199} This method is simple and cost-effective, and can be widely applied in various environments, including primary hospitals, community clinics, and POC settings.

Moreover, the integration of ML and AI algorithms into 2D nanoparticle-based biosensors promises to revolutionize disease diagnosis by enhancing the accuracy, speed, and reliability of detection. ML algorithms leverage vast datasets of biosensor measurements to learn complex patterns and relationships between input signals and target analytes, enabling robust and adaptive detection algorithms. For example, recent studies have demonstrated the application of ML algorithms for data analysis and pattern recognition in graphene-based biosensors for disease detection, achieving high sensitivity and specificity compared to traditional signal processing techniques.^{200,201} Furthermore, AI-based diagnosis holds potential in overcoming major challenges faced by present biosensing technology, such as sensor drift, background noise, and interferences from complex biological matrices. Similarly, development of advanced models such as the convolutional neural network (CNN) model that could accurately classify the presence of disease biomarkers will revolutionize the detection techniques for *H. pylori* infections. By integrating ML and AI algorithms into biosensor platforms, researchers can develop smart diagnostic systems capable of real-time error correction, self-calibration, and adaptive signal processing, thus improving the reliability and reproducibility of *H. pylori* detection in diverse clinical settings.²⁰² Additionally, the use of AI-based diagnosis opens new avenues for predictive analytics, disease modeling, and precision medicine, enabling proactive management of *H. pylori* infection and its associated complications.

Beyond *H. pylori* detection, these future technologies hold promise for addressing a wide range of healthcare challenges such as other infectious pathogens, cancer biomarkers, environmental pollutants, chronic conditions, personalized medicine, and biological warfare agents. For example, multiplexed biosensors and POC devices can be adapted to detect other infectious agents, such as bacteria, viruses (SARS-CoV-2 antigens), and parasites, enabling rapid diagnosis and surveillance

of emerging infectious diseases.^{173,203–206} Wearable and implantable biosensors offer opportunities for continuous monitoring of biomarkers associated with chronic conditions, such as diabetes, cardiovascular disease, and cancer, enabling early detection of disease progression and timely intervention. Moreover, the integration of 2D nanomaterials with AI amplifies the potential of telemedicine. AI-powered systems optimize the release of therapeutic agents in drug delivery systems and improve the precision of disease detection through data analysis and modeling. Real-time data analysis allows for personalized healthcare by analyzing individual patient data, predicting disease risk, and optimizing treatment strategies based on patient-specific profiles, enabling immediate adjustments and tailored interventions. This convergence of 2D nanomaterials and AI fosters dynamic synergy, offering the prospect of elevated patient outcomes, reduced side effects, and the expedited development of cutting-edge healthcare solutions. Collectively, these future technologies have the potential to transform healthcare delivery, improve patient outcomes, and mitigate the global burden of disease, making them indispensable tools in the fight against *H. pylori* and beyond. Addressing the challenges associated with 2D nanoparticle-based biosensors is crucial for their widespread acceptance and commercialization as frontline diagnostic tools during public health crises. Commercialization of 2D-nanostructure-based biomedical products, as of today, has been lagging in its technologic development owing to multifaceted problems, not exclusively monetary or managerial but more of fundamental, including manufacturing at large scale (with), quality control (with precision in size and shape), and biosafety issues. However, overcoming these challenges requires interdisciplinary collaborations between materials scientists, engineers, bio-scientists and healthcare professionals to ensure the development of practical and deployable biosensing solutions tailored to the needs of epidemic management. Additionally, regulatory frameworks and quality assurance protocols must be established to guarantee the safety, efficacy, and reliability of 2D nanoparticle-based biosensors for use in clinical settings.

Data availability

No data was used for the research described in the article.

Conflicts of interest

There are no conflicts to declare.

References

- 1 R. Bückler, M. Azevedo-Vethacke, C. Groll, D. Garten, C. Josenhans, S. Suerbaum and S. Schreiber, Helicobacter pylori colonization critically depends on postprandial gastric conditions, *Sci. Rep.*, 2012, 2, 994.
- 2 C. S. Goodwin, R. K. McCulloch, J. A. Armstrong and S. H. Wee, Unusual Cellular Fatty Acids and Distinctive Ultrastructure in a New Spiral Bacterium (Campylobacter



- Pyloridis) from the Human Gastric Mucosa, *J. Med. Microbiol.*, 1985, **19**, 257–267.
- 3 D. M. Jones, A. Curry and A. J. Fox, An Ultrastructural Study of the Gastric Campylobacter-like Organism ‘Campylobacter pyloridis’, *Microbiology*, 1985, **131**, 2335–2341.
 - 4 S. Moens and J. Vanderleyden, Functions of Bacterial Flagella, *Crit. Rev. Microbiol.*, 1996, **22**, 67–100.
 - 5 S. Hirukawa, H. Sagara, S. Kaneto, T. Kondo, K. Kiga, T. Sanada, H. Kiyono and H. Mimuro, Characterization of morphological conversion of *Helicobacter pylori* under anaerobic conditions, *Microbiol. Immunol.*, 2018, **62**, 221–228.
 - 6 J. K. Y. Hooi, W. Y. Lai, W. K. Ng, M. M. Y. Suen, F. E. Underwood, D. Tanyingoh, P. Malfertheiner, D. Y. Graham, V. W. S. Wong, J. C. Y. Wu, F. K. L. Chan, J. J. Y. Sung, G. G. Kaplan and S. C. Ng, Global Prevalence of *Helicobacter pylori* Infection: Systematic Review and Meta-Analysis, *Gastroenterology*, 2017, **153**, 420–429.
 - 7 Y. Li, H. Choi, K. Leung, F. Jiang, D. Y. Graham and W. K. Leung, Global prevalence of *Helicobacter pylori* infection between 1980 and 2022: a systematic review and meta-analysis, *Lancet Gastroenterol. Hepatol.*, 2023, **8**, 553–564.
 - 8 L. E. Wroblewski, R. M. Peek and K. T. Wilson, *Helicobacter pylori* and Gastric Cancer: Factors That Modulate Disease Risk, *Clin. Microbiol. Rev.*, 2010, **23**, 713–739.
 - 9 N. Uemura, S. Okamoto, S. Yamamoto, N. Matsumura, S. Yamaguchi, M. Yamakido, K. Taniyama, N. Sasaki and R. J. Schlemper, *Helicobacter pylori* infection and the development of gastric cancer, *N. Engl. J. Med.*, 2001, **345**, 784.
 - 10 C. Dunne, Factors that mediate colonization of the human stomach by *Helicobacter pylori*, *World J. Gastroenterol.*, 2014, **20**, 5610.
 - 11 B. M. Roesler, E. M. A. Rabelo-Gonçalves and J. M. R. Zeitune, Virulence Factors of *Helicobacter pylori*: A Review, *Clin. Med. Insights: Gastroenterol.*, 2014, **7**, S13760.
 - 12 N. Uemura, S. Okamoto, S. Yamamoto, N. Matsumura, S. Yamaguchi, M. Yamakido, K. Taniyama, N. Sasaki and R. J. Schlemper, *Helicobacter pylori* infection and the development of gastric cancer, *N. Engl. J. Med.*, 2001, **345**, 784.
 - 13 S. C. Shah, M. B. Piazuelo, E. J. Kuipers and D. Li, AGA Clinical Practice Update on the Diagnosis and Management of Atrophic Gastritis: Expert Review, *Gastroenterology*, 2021, **161**, 1325–1332.
 - 14 B. Guevara and A. G. Cogdill, *Helicobacter pylori*: A Review of Current Diagnostic and Management Strategies, *Dig. Dis. Sci.*, 2020, **65**, 1917–1931.
 - 15 P. Malfertheiner, M. C. Camargo, E. El-Omar, J.-M. Liou, R. Peek, C. Schulz, S. I. Smith and S. Suerbaum, *Helicobacter pylori* infection, *Nat. Rev. Dis. Prim.*, 2023, **9**, 19.
 - 16 M. Ferwana, Accuracy of urea breath test in *Helicobacter pylori* infection: Meta-analysis, *World J. Gastroenterol.*, 2015, **21**, 1305.
 - 17 O. C. Aktepe, Five methods for detection of *Helicobacter pylori* in the Turkish population, *World J. Gastroenterol.*, 2011, **17**, 5172.
 - 18 D. T. Quach, R. Aoki, A. Iga, Q. D. Le, T. Kawamura, K. Yamashita, S. Tanaka, M. Yoshihara and T. Hiyama, Diagnostic Accuracy of *H. pylori* Status by Conventional Endoscopy: Time-Trend Change After Eradication and Impact of Endoscopic Image Quality, *Front. Med.*, 2022, **8**, 830730.
 - 19 M. P. Dore and G. M. Pes, What Is New in *Helicobacter pylori* Diagnosis. An Overview, *J. Clin. Med.*, 2021, **10**, 2091.
 - 20 D. S. Bordin, I. N. Voynovan, D. N. Andreev and I. V. Maev, Current *Helicobacter pylori* Diagnostics, *Diagnostics*, 2021, **11**, 1458.
 - 21 J. Y. Lee and N. Kim, Diagnosis of *Helicobacter pylori* by invasive test: histology, *Ann. Transl. Med.*, 2015, **3**, 10.
 - 22 B. Dai, A. Xiang, D. Qu, G. Chen, L. Wang, W. Wang, D. Zhai, L. Wang and Z. Lu, Rapid and Sensitive Assay of *Helicobacter pylori* With One-Tube RPA-CRISPR/Cas12 by Portable Array Detector for Visible Analysis of Thermo-static Nucleic Acid Amplification, *Front. Microbiol.*, 2022, **13**, 858247.
 - 23 M. M. Ahmad, D. S. Ahmed, A. H. M. Rowshon, S. C. Dhar, M. Rahman, M. Hasan, C. Beglinger, N. Gyr and A. K. A. Khan, Long-Term Re-Infection Rate after *Helicobacter pylori* Eradication in Bangladeshi Adults, *Digestion*, 2007, **75**, 173–176.
 - 24 L. P. Andersen and T. Wadström, *Helicobacter pylori: Physiology and Genetics*, 2001.
 - 25 N. Albertson, I. Wenngren and J. E. Sjöström, Growth and survival of *Helicobacter pylori* in defined medium and susceptibility to Brij 78, *J. Clin. Microbiol.*, 1998, **36**, 1232.
 - 26 A. Shiotani, M. P. Dore and D. Y. Graham, *Helicobacter pylori*, Springer, Japan, Tokyo, 2016, pp. 143–155.
 - 27 Y. Kohli, T. Kato, S. Ito, H. Miyazi, T. Azuma, K. Nagata, H. Tsuruta, S. Matsui, K. Oka and M. Nakamura, Diagnostic System for *Helicobacter pylori* Urease Based on a pHsensitive Field Effect Transistor, *Dig. Endosc.*, 1995, **7**, 27–34.
 - 28 A. B. Adamopoulos, G. S. Stergiou, G. N. Sakizlis, D. G. Tiniakos, E. G. Nasothimiou, D. K. Sioutis and A. D. Achimastos, Diagnostic value of rapid urease test and urea breath test for *Helicobacter pylori* detection in patients with Billroth II gastrectomy: a prospective controlled trial, *Dig. Liver Dis.*, 2009, **41**, 4–8.
 - 29 N. Vakil and D. Vaira, Non-invasive tests for the diagnosis of *H. pylori* infection, *Rev. Gastroenterol. Disord.*, 2004, **4**, 1–6.
 - 30 G. Oderda, Detection of *Helicobacter pylori* in stool specimens by non-invasive antigen enzyme immunoassay in children: multicentre Italian study, *BMJ*, 2000, **320**, 347–348.
 - 31 X. Zhou, J. Su, G. Xu and G. Zhang, Accuracy of stool antigen test for the diagnosis of *Helicobacter pylori* infection in children: A meta-analysis, *Clin. Res. Hepatol. Gastroenterol.*, 2014, **38**, 629–638.
 - 32 R. A. Hussein, M. T. S. Al-Ouqaili and Y. H. Majeed, Detection of *Helicobacter Pylori* infection by invasive and non-invasive techniques in patients with gastrointestinal diseases from Iraq: A validation study, *PLoS One*, 2021, **16**, e0256393.



- 33 Z. Rafiei-Sarmazdeh, S. Morteza Zahedi-Dizaji and A. Kafi Kang, *Nanostructures*, IntechOpen, 2020.
- 34 T. Hu, X. Mei, Y. Wang, X. Weng, R. Liang and M. Wei, Two-dimensional nanomaterials: fascinating materials in biomedical field, *Sci. Bull.*, 2019, **64**, 1707–1727.
- 35 V. K. Vaishnav, K. Prasad, R. Yadav, A. Aharwar and B. Nath Tiwary, *Recent Advances in Biosensor Technology*, Bentham Science Publishers, 2023, pp. 45–77.
- 36 D. Y. Graham and M. Miftahussurur, Helicobacter pylori urease for diagnosis of Helicobacter pylori infection: A mini review, *J. Adv. Res.*, 2018, **13**, 51–57.
- 37 V. Cussac, R. L. Ferrero and A. Labigne, Expression of Helicobacter pylori urease genes in Escherichia coli grown under nitrogen-limiting conditions, *J. Bacteriol.*, 1992, **174**, 2466–2473.
- 38 A. Labigne, V. Cussac and P. Courcoux, Shuttle cloning and nucleotide sequences of Helicobacter pylori genes responsible for urease activity, *J. Bacteriol.*, 1991, **173**, 1920–1931.
- 39 H. C. Sharndama and I. E. Mba, Helicobacter pylori: an up-to-date overview on the virulence and pathogenesis mechanisms, *Brazilian J. Microbiol.*, 2022, **53**, 33–50.
- 40 M. Sugimoto and Y. Yamaoka, The association of vacA genotype and Helicobacter pylori-related disease in Latin American and African populations, *Clin. Microbiol. Infect.*, 2009, **15**, 835–842.
- 41 M. Karbalaeei, A. Talebi Bezmin Abadi and M. Keikha, Clinical relevance of the cagA and vacA s1m1 status and antibiotic resistance in Helicobacter pylori: a systematic review and meta-analysis, *BMC Infect. Dis.*, 2022, **22**, 573.
- 42 M. Hatakeyama, Structure and function of Helicobacter pylori CagA, the first-identified bacterial protein involved in human cancer, *Proc. Jpn. Acad., Ser. B*, 2017, **93**, 196–219.
- 43 L. Jeyamani, J. Jayarajan, V. Leelakrishnan and M. Swaminathan, CagA and VacA genes of Helicobacter pylori and their clinical relevance, *Indian J. Pathol. Microbiol.*, 2018, **61**, 66.
- 44 T. L. Cover, D. B. Lacy and M. D. Ohi, The Helicobacter pylori Cag Type IV Secretion System, *Trends Microbiol.*, 2020, **28**, 682–695.
- 45 J. M. Noto and R. M. Peek, The *Helicobacter pylori* cag Pathogenicity Island, *Methods Mol. Biol.*, 2012, **921**, 41–50.
- 46 K. Saxena, B. T. Murti, P.-K. Yang, B. D. Malhotra, N. Chauhan and U. Jain, Fabrication of a Molecularly Imprinted Nano-Interface-Based Electrochemical Biosensor for the Detection of CagA Virulence Factors of H. pylori, *Biosensors*, 2022, **12**, 1066.
- 47 J. S. Vital, L. Tanoeiro, R. Lopes-Oliveira and F. F. Vale, Biomarker Characterization and Prediction of Virulence and Antibiotic Resistance from Helicobacter pylori Next Generation Sequencing Data, *Biomolecules*, 2022, **12**, 691.
- 48 M. Oleastro, R. Cordeiro, A. Menard and J. P. Gomes, Allelic Diversity among Helicobacter pylori Outer Membrane Protein Genes homB and homA Generated by Recombination, *J. Bacteriol.*, 2010, **192**, 3961–3968.
- 49 X. Huang, D. Zhaomin, Q. Zhang, W. Li, B. Wang and M. Li, Relationship between the iceA gene of Helicobacter pylori and clinical outcomes, *Ther. Clin. Risk Manag.*, 2016, **12**, 1085–1092.
- 50 Y.-L. Su, H.-L. Huang, B.-S. Huang, P.-C. Chen, C.-S. Chen, H.-L. Wang, P.-H. Lin, M.-S. Chieh, J.-J. Wu, J.-C. Yang and L.-P. Chow, Combination of OipA, BabA, and SabA as candidate biomarkers for predicting Helicobacter pylori-related gastric cancer, *Sci. Rep.*, 2016, **6**, 36442.
- 51 C. Anichini, W. Czepa, D. Pakulski, A. Aliprandi, A. Ciesielski and P. Samori, Chemical sensing with 2D materials, *Chem. Soc. Rev.*, 2018, **47**, 4860–4908.
- 52 J. Li, H. Zeng, Z. Zeng, Y. Zeng and T. Xie, Promising graphene-based nanomaterials and their biomedical applications and potential risks: A comprehensive review, *ACS Biomater. Sci. Eng.*, 2021, **7**, 5363–5396.
- 53 J. Wang, L. Sui, J. Huang, L. Miao, Y. Nie, K. Wang, Z. Yang, Q. Huang, X. Gong, Y. Nan and K. Ai, MoS₂-based nanocomposites for cancer diagnosis and therapy, *Bioact. Mater.*, 2021, **6**, 4209–4242.
- 54 J. Wang, G. Li and L. Li, *Two-dimensional Materials – Synthesis, Characterization and Potential Applications*, InTech, 2016.
- 55 Y. Chen, Z. Fan, Z. Zhang, W. Niu, C. Li, N. Yang, B. Chen and H. Zhang, Two-Dimensional Metal Nanomaterials: Synthesis, Properties, and Applications, *Chem. Rev.*, 2018, **118**, 6409–6455.
- 56 P. Nguyen Tri, C. Ouellet-Plamondon, S. Rtimi, A. A. Assadi and T. A. Nguyen, *Noble Metal-Metal Oxide Hybrid Nanoparticles*, Elsevier, 2019, pp. 51–63.
- 57 G. Habibullah, J. Viktorova and T. Ruml, Current Strategies for Noble Metal Nanoparticle Synthesis, *Nanoscale Res. Lett.*, 2021, **16**, 47.
- 58 R. Indiarto, L. P. A. Indriana, R. Andoyo, E. Subroto and B. Nurhadi, Bottom-up nanoparticle synthesis: a review of techniques, polyphenol-based core materials, and their properties, *Eur. Food Res. Technol.*, 2022, **248**, 1–24.
- 59 A. Tripathi and T. B. Pirzadah, *Synthesis of Bionanomaterials for Biomedical Applications*, Elsevier, 2023, pp. 57–76.
- 60 T. Lage, R. O. Rodrigues, S. Catarino, J. Gallo, M. Bañobre-López and G. Minas, Graphene-Based Magnetic Nanoparticles for Theranostics: An Overview for Their Potential in Clinical Application, *Nanomaterials*, 2021, **11**, 1073.
- 61 Z. Li, X. Li, M. Jian, G. S. Geleta and Z. Wang, Two-Dimensional Layered Nanomaterial-Based Electrochemical Biosensors for Detecting Microbial Toxins, *Toxins*, 2019, **12**, 20.
- 62 G. Sanità, B. Carrese and A. Lamberti, Nanoparticle Surface Functionalization: How to Improve Biocompatibility and Cellular Internalization, *Front. Mol. Biosci.*, 2020, **7**, 587012.
- 63 A. Shakeri, S. Khan and T. F. Didar, Conventional and emerging strategies for the fabrication and functionalization of PDMS-based microfluidic devices, *Lab Chip*, 2021, **21**, 3053–3075.
- 64 W. Putzbach and N. Ronkainen, Immobilization Techniques in the Fabrication of Nanomaterial-Based Electrochemical Biosensors: A Review, *Sensors*, 2013, **13**, 4811–4840.
- 65 Y. Zhou, Y. Fang and R. Ramasamy, Non-Covalent Functionalization of Carbon Nanotubes for Electrochemical Biosensor Development, *Sensors*, 2019, **19**, 392.



- 66 E. B. Aydin, M. Aydin and M. K. Sezginurk, *New Developments in Nanosensors for Pharmaceutical Analysis*, Elsevier, 2019, pp. 47–78.
- 67 D. Kala, T. K. Sharma, S. Gupta, V. Verma, A. Thakur, A. Kaushal, A. V. Trukhanov and S. V. Trukhanov, Graphene Oxide Nanoparticles Modified Paper Electrode as a Biosensing Platform for Detection of the htrA Gene of *O. tsutsugamushi*, *Sensors*, 2021, **21**, 4366.
- 68 F. Löscher, T. Ruckstuhl, T. Jaworek, G. Wegner and S. Seeger, Immobilization of Biomolecules on Langmuir–Blodgett Films of Regenerative Cellulose Derivatives, *Langmuir*, 1998, **14**, 2786–2789.
- 69 A. Sassolas, L. J. Blum and B. D. Leca-Bouvier, Immobilization strategies to develop enzymatic biosensors, *Biotechnol. Adv.*, 2012, **30**, 489–511.
- 70 S. Khan, B. Burciu, C. D. M. Filipe, Y. Li, K. Dellinger and T. F. Didar, DNAzyme-Based Biosensors: Immobilization Strategies, Applications, and Future Prospective, *ACS Nano*, 2021, **15**, 13943–13969.
- 71 V. Kathiresan, D. Thirumalai, T. Rajarathinam, M. Yeom, J. Lee, S. Kim, J.-H. Yoon and S.-C. Chang, A simple one-step electrochemical deposition of bioinspired nanocomposite for the non-enzymatic detection of dopamine, *J. Anal. Sci. Technol.*, 2021, **12**, 5.
- 72 R. Ahmad, O. S. Wolfbeis, Y.-B. Hahn, H. N. Alshareef, L. Torsi and K. N. Salama, Deposition of nanomaterials: A crucial step in biosensor fabrication, *Mater. Today Commun.*, 2018, **17**, 289–321.
- 73 S. Palani, J. P. Kenison, S. Sabuncu, T. Huang, F. Civitci, S. Esener and X. Nan, Multispectral Localized Surface Plasmon Resonance (msLSPR) Reveals and Overcomes Spectral and Sensing Heterogeneities of Single Gold Nanoparticles, *ACS Nano*, 2023, **17**, 2266–2278.
- 74 V. Kumar, S. Mishra, R. Sharma, J. Agarwal, U. Ghoshal, T. Khanna, L. K. Sharma, S. K. Verma, P. Mishra and S. Tiwari, Development of RNA-Based Assay for Rapid Detection of SARS-CoV-2 in Clinical Samples, *Intervirology*, 2022, **65**, 181–187.
- 75 Y. Huang, X. Dong, Y. Liu, L.-J. Li and P. Chen, Graphene-based biosensors for detection of bacteria and their metabolic activities, *J. Mater. Chem.*, 2011, **21**, 12358.
- 76 D. Kwong Hong Tsang, T. J. Lieberthal, C. Watts, I. E. Dunlop, S. Ramadan, A. E. del Rio Hernandez and N. Klein, Chemically Functionalised Graphene FET Biosensor for the Label-free Sensing of Exosomes, *Sci. Rep.*, 2019, **9**, 13946.
- 77 L. Chen, G. Li, A. Yang, J. Wu, F. Yan and H. Ju, A DNA-functionalized graphene field-effect transistor for quantitation of vascular endothelial growth factor, *Sensors Actuators B Chem.*, 2022, **351**, 130964.
- 78 H. Du, J. Ye, J. Zhang, X. Huang and C. Yu, Graphene Nanosheets Modified Glassy Carbon Electrode as a Highly Sensitive and Selective Voltammetric Sensor for Rutin, *Electroanalysis*, 2010, **22**, 2399–2406.
- 79 A. M. Hareedy, S. M. Derayea, A. A. Gahlan, M. A. Omar and G. A. Saleh, Graphene oxide modified glassy carbon electrode for determination of linagliptin in dosage form, biological fluids, and rats' feces using square wave voltammetry, *Arab. J. Chem.*, 2022, **15**, 103663.
- 80 S. Bagherzadeh-Nobari and R. Kalantarinejad, Real-time label-free detection of DNA hybridization using a functionalized graphene field effect transistor: a theoretical study, *J. Nanoparticle Res.*, 2021, **23**, 185.
- 81 J. Yamaguchi, H. Hayashi, H. Jippo, A. Shiotari, M. Ohtomo, M. Sakakura, N. Hieda, N. Aratani, M. Ohfuchi, Y. Sugimoto, H. Yamada and S. Sato, Small bandgap in atomically precise 17-atom-wide armchair-edged graphene nanoribbons, *Commun. Mater.*, 2020, **1**, 36.
- 82 H. Chen, B. Luo, S. Wu, S. Shi, Q. Dai, Z. Peng and M. Zhao, Microfluidic Biosensor Based on Molybdenum Disulfide (MoS₂) Modified Thin-Core Microfiber for Immune Detection of *Toxoplasma gondii*, *Sensors*, 2023, **23**, 5218.
- 83 A. K. Bhatia and S. Dewangan, *Handbook of Biomolecules*, Elsevier, 2023, pp. 259–274.
- 84 A. Asadzadeh-Firouzabadi, H. R. Zare and N. Nasirizadeh, Electrochemical Biosensor for Detection of Target DNA Sequence and Single-Base Mismatch Related to *Helicobacter Pylori* Using Chlorogenic Acid as Hybridization Indicator, *J. Electrochem. Soc.*, 2016, **163**, B43–B48.
- 85 S. Gupta, U. Jain, B. T. Murti, A. D. Putri, A. Tiwari and N. Chauhan, Nanohybrid-based immunosensor prepared for *Helicobacter pylori* BabA antigen detection through immobilized antibody assembly with @ Pdnano/rGO/PEDOT sensing platform, *Sci. Rep.*, 2020, **10**, 21217.
- 86 K. Saxena, N. Chauhan, B. D. Malhotra and U. Jain, A molecularly imprinted polymer-based electrochemical biosensor for detection of VacA virulence factor of *H. pylori* causing gastric cancer, *Process Biochem.*, 2023, **130**, 87–95.
- 87 H. Jaradat, B. M. Hryniewicz, I. A. Pašti, T. L. Valério, A. Al-Hamry, L. F. Marchesi, M. Vidotti and O. Kanoun, Detection of *H. pylori* outer membrane protein (HopQ) biomarker using electrochemical impedimetric immunosensor with polypyrrole nanotubes and carbon nanotubes nanocomposite on screen-printed carbon electrode, *Biosens. Bioelectron.*, 2024, **249**, 115937.
- 88 T. B. A. Akib, S. Mostufa, M. M. Rana, M. B. Hossain and M. R. Islam, A performance comparison of heterostructure surface plasmon resonance biosensor for the diagnosis of novel coronavirus SARS-CoV-2, *Opt. Quantum Electron.*, 2023, **55**, 448.
- 89 T. N. Ghosh, D. R. Rotake and S. G. Singh, 2D vanadium disulfide nanosheets assisted ultrasensitive, rapid, and label-free electrochemical quantification of cancer biomarker (MMP-2), *Nanotechnology*, 2023, **34**, 395501.
- 90 C. Rodriguez-Emmenegger, E. Brynda, T. Riedel, M. Houska, V. Šubr, A. B. Alles, E. Hasan, J. E. Gautrot and W. T. S. Huck, Polymer Brushes Showing Non-Fouling in Blood Plasma Challenge the Currently Accepted Design of Protein Resistant Surfaces, *Macromol. Rapid Commun.*, 2011, **32**, 952–957.
- 91 H. Chen, Z. Wang, Z. Liu, Q. Niu, X. Wang, Z. Miao, H. Zhang, J. Wei, M. Wan and C. Mao, Construction of



- 3D electrochemical cytosensor by layer-by-layer assembly for ultra-sensitive detection of cancer cells, *Sensors Actuators B Chem.*, 2021, **329**, 128995.
- 92 R. Wang, H. You, Y. Zhang, Z. Li, Y. Ding, Q. Qin, H. Wang, J. Sun, Y. Jia, F. Liu, J. Ma and X. Cheng, Constructing (reduced) graphene oxide enhanced polypyrrole/ceramic composite membranes for water remediation, *J. Memb. Sci.*, 2022, **659**, 120815.
- 93 W. Zhang, Z. Tian, S. Yang, J. Rich, S. Zhao, M. Klingeborn, P.-H. Huang, Z. Li, A. Stout, Q. Murphy, E. Patz, S. Zhang, G. Liu and T. J. Huang, Electrochemical micro-aptasensors for exosome detection based on hybridization chain reaction amplification, *Microsystems Nanoeng.*, 2021, **7**, 63.
- 94 M. T. Hwang, M. Heiranian, Y. Kim, S. You, J. Leem, A. Taqieddin, V. Faramarzi, Y. Jing, I. Park, A. M. van der Zande, S. Nam, N. R. Aluru and R. Bashir, Ultrasensitive detection of nucleic acids using deformed graphene channel field effect biosensors, *Nat. Commun.*, 2020, **11**, 1543.
- 95 J. Wang, D. Chen, W. Huang, N. Yang, Q. Yuan and Y. Yang, Aptamer-functionalized field-effect transistor biosensors for disease diagnosis and environmental monitoring, *Exploration*, 2023, **3**, 20210027.
- 96 L. Wang, K. Cui, P. Wang, M. Pei and W. Guo, A sensitive electrochemical DNA sensor for detecting *Helicobacter pylori* based on accordion-like Ti3C2Tx: a simple strategy, *Anal. Bioanal. Chem.*, 2021, **413**, 4353–4362.
- 97 H. Zhao and Y. Lei, 3D Nanostructures for the Next Generation of High-Performance Nanodevices for Electrochemical Energy Conversion and Storage, *Adv. Energy Mater.*, 2020, **10**(28), 2001460.
- 98 M. Byakodi, N. S. Shrikrishna, R. Sharma, S. Bhansali, Y. Mishra, A. Kaushik and S. Gandhi, Emerging 0D, 1D, 2D, and 3D nanostructures for efficient point-of-care biosensing, *Biosens. Bioelectron. X*, 2022, **12**, 100284.
- 99 E. Qiu, Z. Li and S. Han, Methods for detection of *Helicobacter pylori* from stool sample: current options and developments, *Brazilian J. Microbiol.*, 2021, **52**, 2057–2062.
- 100 S. Su and P. M. Kang, Systemic Review of Biodegradable Nanomaterials in Nanomedicine, *Nanomaterials*, 2020, **10**, 656.
- 101 R. Vainionpää and P. Leinikki, *Encyclopedia of Virology*, Elsevier, 2008, pp. 29–37.
- 102 C. Fernandes, M. Jathar, B. K. S. Sawant and T. Warde, Scale-Up of Nanoparticle Manufacturing Process, *Pharmaceutical Process Engineering and Scale-up Principles*, 2023, vol. 13, pp. 173–203.
- 103 H. Zhu, H. Zhang, Y. Xu, S. Laššáková, M. Korabečná and P. Neuzil, PCR past, present and future, *Biotechniques*, 2020, **69**, 317–325.
- 104 E. Andresen, F. Islam, C. Prinz, P. Gehrman, K. Licha, J. Roik, S. Recknagel and U. Resch-Genger, Assessing the reproducibility and up-scaling of the synthesis of Er,Yb-doped NaYF4-based upconverting nanoparticles and control of size, morphology, and optical properties, *Sci. Rep.*, 2023, **13**, 2288.
- 105 L. Gautier, Reproducibility in the lab: the proof is in the protocol, *Biotechniques*, 2022, **73**, 71–74.
- 106 J. P. Gallagher, P. A. Twohig, A. Crnic and F. A. Rochling, Illicit Drug Use and Endoscopy: When Do We Say No?, *Dig. Dis. Sci.*, 2022, **67**, 5371–5381.
- 107 M. Khalid and M. Abdollahi, Toxicity of inorganic nanoparticles, *Compr. Anal. Chem.*, 2022, **99**, 25–85.
- 108 G. Chakraborty, I.-H. Park, R. Medishetty and J. J. Vittal, Two-Dimensional Metal-Organic Framework Materials: Synthesis, Structures, Properties and Applications, *Chem. Rev.*, 2021, **121**, 3751–3891.
- 109 T. R. Kyriakides, A. Raj, T. H. Tseng, H. Xiao, R. Nguyen, F. S. Mohammed, S. Halder, M. Xu, M. J. Wu, S. Bao and W. C. Sheu, Biocompatibility of nanomaterials and their immunological properties, *Biomed. Mater.*, 2021, **16**, 042005.
- 110 S. S. Mirzaei, N. Mehrdadi, M. Pourmadadi, M. Ahmadi and S. Meknatkhan, Novel detection of *H. pylori* using ultrasensitive electrochemical aptasensor based on surface modified graphene oxide doped gold nanoparticles conjugated polythiophene, *Microchem. J.*, 2024, **200**, 110279.
- 111 B. Cai, Z. Xia, J. Wang, S. Wu and X. Jin, Reduced Graphene Oxide-Based Field Effect Transistor Biosensors for High-Sensitivity miRNA21 Detection, *ACS Appl. Nano Mater.*, 2022, **5**, 12035–12044.
- 112 Y. Yim, H. Shin, S. M. Ahn and D.-H. Min, Graphene oxide-based fluorescent biosensors and their biomedical applications in diagnosis and drug discovery, *Chem. Commun.*, 2021, **57**, 9820–9833.
- 113 A. M. Salama, G. Yasin, M. Zourob and J. Lu, Fluorescent Biosensors for the Detection of Viruses Using Graphene and Two-Dimensional Carbon Nanomaterials, *Biosensors*, 2022, **12**, 460.
- 114 T. Yang and T. V. Duncan, Challenges and potential solutions for nanosensors intended for use with foods, *Nat. Nanotechnol.*, 2021, **16**, 251–265.
- 115 S. Yamamoto, A. Tanaka, S. Kobayashi, Y. Oshiro, M. Ozeki, K. Maeda, K. Matsuda, K. Miyo, T. Mizoue, W. Sugiura, H. Mitsuya, H. Sugiyama and N. Ohmagari, Consistency of the results of rapid serological tests for SARS-CoV-2 among healthcare workers in a large national hospital in Tokyo, Japan, *Glob. Heal. Med.*, 2021, **3**, 90–94.
- 116 O. Yilmaz, N. Sen, A. A. Küpelioglu and I. Simsek, Detection of *H. pylori* infection by ELISA and western blot techniques and evaluation of anti CagA seropositivity in adult Turkish dyspeptic patients, *World J. Gastroenterol.*, 2006, **12**, 5375.
- 117 K. Saxena, A. Kumar, N. Chauhan, M. Khanuja, B. D. Malhotra and U. Jain, Electrochemical Immunosensor for Detection of *H. pylori* Secretory Protein VacA on g-C 3 N 4/ZnO Nanocomposite-Modified Au Electrode, *ACS Omega*, 2022, **7**, 32292–32301.
- 118 E. Janik-Karpinska, M. Ceremuga, M. Niemcewicz, M. Podogrocki, M. Stela, N. Cichon and M. Bijak, Immunosensors—The Future of Pathogen Real-Time Detection, *Sensors*, 2022, **22**, 9757.
- 119 S. Kizhepat, A. S. Rasal, J.-Y. Chang and H.-F. Wu, Development of Two-Dimensional Functional Nanomaterials for Biosensor Applications: Opportunities, Challenges, and Future Prospects, *Nanomaterial*, 2023, **13**, 1520.



- 120 I. S. Raja, M. Vedhanayagam, D. R. Preeth, C. Kim, J. H. Lee and D. W. Han, Development of Two-Dimensional Nanomaterials Based Electrochemical Biosensors on Enhancing the Analysis of Food Toxicants, *Int. J. Mol. Sci.*, 2021, **22**, 3277.
- 121 R. Khan, A. Radoi, S. Rashid, A. Hayat, A. Vasilescu and S. Andreescu, Two-Dimensional Nanostructures for Electrochemical Biosensor, *Sensors*, 2021, **21**, 3369.
- 122 R. Hao, L. Liu, J. Yuan, L. Wu and S. Lei, Recent Advances in Field Effect Transistor Biosensors: Designing Strategies and Applications for Sensitive Assay, *Biosensors*, 2023, **13**, 426.
- 123 X. Su and S. F. Li, Serological determination of *Helicobacter pylori* infection using sandwiched and enzymatically amplified piezoelectric biosensor, *Anal. Chim. Acta*, 2001, **429**, 27–36.
- 124 A. Kumar, H. K. Rathore, D. Sarkar and A. Shukla, Nanoarchitected transition metal oxides and their composites for supercapacitors, *Electrochem. Sci. Adv.*, 2022, **2**, e2100187.
- 125 L. M. H. Al-Ameri, M. L. Israa and Saqari, The Optical Biosensor was used to Determine Antibiotic Micsagainst *H. pylori*, *J. Res. Med. Dent. Sci.*, 2022, **10**, 275–280.
- 126 P. M. Arun Kumar Gupta, Ashwini Yadav and Parismita Koch, *Biosensors in Food Safety and Quality*, 1st edn, 2022, p. 10.
- 127 B. Dick and M. Jones, The commercialization of molecular biology: Walter Gilbert and the Biogen startup, *Hist. Technol.*, 2017, **33**, 126–151.
- 128 F. Gong, H. Wei, Q. Li, L. Liu and B. Li, Evaluation and Comparison of Serological Methods for COVID-19 Diagnosis, *Front. Mol. Biosci.*, 2021, **8**, 682405.
- 129 S. Malik, J. Singh, R. Goyat, Y. Saharan, V. Chaudhry, A. Umar, A. A. Ibrahim, S. Akbar, S. Ameen and S. Baskoutas, Nanomaterials-based biosensor and their applications: A review, *Heliyon*, 2023, **9**, e19929.
- 130 R. Abdel-Karim, Nanotechnology-Enabled Biosensors: A Review of Fundamentals, Materials, Applications, Challenges, and Future Scope, *Biomed. Mater. Dev.*, 2024, **2**, 759–777.
- 131 A. Seitak, S. Luo, N. Cai, K. Liao, A.-M. Pappa, S. Lee and V. Chan, Emergence of MXene-based electrochemical biosensors for biomolecule and pathogen detection, *Sensors and Actuators Reports*, 2023, **6**, 100175.
- 132 D. Hristov, C. Rodriguez-Quijada, J. Gomez-Marquez and K. Hamad-Schifferli, Designing Paper-Based Immunoassays for Biomedical Applications, *Sensors*, 2019, **19**, 554.
- 133 S. A. Hashemi, S. M. Mousavi, S. Bahrani and S. Ramakrishna, Integrated polyaniline with graphene oxide-iron tungsten nitride nanoflakes as ultrasensitive electrochemical sensor for precise detection of 4-nitrophenol within aquatic media, *J. Electroanal. Chem.*, 2020, **873**, 114406.
- 134 L. Chen, E. Hwang and J. Zhang, Fluorescent Nanobiosensors for Sensing Glucose, *Sensors*, 2018, **18**, 1440.
- 135 S. Senapati, A. K. Mahanta, S. Kumar and P. Maiti, Controlled drug delivery vehicles for cancer treatment and their performance, *Signal Transduct. Target. Ther.*, 2018, **3**, 7.
- 136 M. R. Bernsen, M. van Straten, G. Kotek, E. A. H. Warnert, J. C. Haeck, A. Ruggiero, P. A. Wielopolski and G. P. Krestin, Computed Tomography and Magnetic Resonance Imaging, *Recent Results Cancer Res.*, 2020, **216**, 31–110.
- 137 V. Srimaneepong, H. E. Skallevoid, Z. Khurshid, M. S. Zafar, D. Rokaya and J. Sapkota, Graphene for Antimicrobial and Coating Application, *Int. J. Mol. Sci.*, 2022, **23**, 499.
- 138 P. J. Vikesland, Nanosensors for water quality monitoring, *Nat. Nanotechnol.*, 2018, **13**, 651–660.
- 139 S. Dhall, B. R. Mehta, A. K. Tyagi and K. Sood, A review on environmental gas sensors: Materials and technologies, *Sensors Int.*, 2021, **2**, 100116.
- 140 Z. Wang and B. Mi, Environmental Applications of 2D Molybdenum Disulfide (MoS₂) Nanosheets, *Environ. Sci. Technol.*, 2017, **51**, 8229–8244.
- 141 A. Aghababai Beni and H. Jabbari, Nanomaterials for Environmental Applications, *Results Eng.*, 2022, **15**, 100467.
- 142 J. Chen, Y. Guo, X. Zhang, J. Liu, P. Gong, Z. Su, L. Fan and G. Li, Emerging Nanoparticles in Food: Sources, Application, and Safety, *J. Agric. Food Chem.*, 2023, **71**, 3564–3582.
- 143 R. Biswas, M. Alam, A. Sarkar, M. I. Haque, M. M. Hasan and M. Hoque, Application of nanotechnology in food: processing, preservation, packaging and safety assessment, *Heliyon*, 2022, **8**, e11795.
- 144 W. Wen, Y. Song, X. Yan, C. Zhu, D. Du, S. Wang, A. M. Asiri and Y. Lin, Recent advances in emerging 2D nanomaterials for biosensing and bioimaging applications, *Mater. Today*, 2018, **21**, 164–177.
- 145 Z. Chen, T. Aziz, H. Sun, A. Ullah, A. Ali, L. Cheng, R. Ullah and F. U. Khan, Advances and Applications of Cellulose Bio-Composites in Biodegradable Materials, *J. Polym. Environ.*, 2023, **31**, 2273–2284.
- 146 S. S. Mirzaei, M. Pourmadadi, A. Foroozandeh, A. A. Moghaddam, M. Soltani, N. Basirhaghighi and M. Ahmadi, Enhancing *H. pylori* detection: ultrasensitive electrochemical aptasensor with Au-doped CQDs and polythiophene conjugation, *J. Appl. Electrochem.*, 2024, **54**, 1887–1900.
- 147 U. Jain, S. Gupta, S. Soni, M. P. Khurana and N. Chauhan, Triple-nanostructuring-based noninvasive electro-immune sensing of CagA toxin for *Helicobacter pylori* detection, *Helicobacter*, 2020, **25**, e12706.
- 148 M. Mathew and C. S. Rout, Electrochemical biosensors based on Ti3C2Tx MXene: future perspectives for on-site analysis, *Curr. Opin. Electrochem.*, 2021, **30**, 100782.
- 149 J. Chao, D. Zhu, Y. Zhang, L. Wang and C. Fan, DNA nanotechnology-enabled biosensors, *Biosens. Bioelectron.*, 2016, **76**, 68–79.
- 150 S. Hajihosseini, N. Nasirizadeh, M. S. Hejazi and P. Yaghmaei, A sensitive DNA biosensor fabricated from gold nanoparticles and graphene oxide on a glassy carbon electrode, *Mater. Sci. Eng. C*, 2016, **61**, 506–515.
- 151 M. V. del Pozo, C. Alonso, F. Pariente and E. Lorenzo, DNA Biosensor for Detection of *Helicobacter pylori* Using Phendione as the Electrochemically Active Ligand in Osmium Complexes, *Anal. Chem.*, 2005, **77**, 2550–2557.



- 152 A. K. Yadav, D. Verma, S. Tabassum and P. R. Solanki, Functionalized Graphitic Carbon Nitride Based Aptasensing Platform for Electrochemical Detection of *Helicobacter Pylori*, *IEEE Sensors Lett.*, 2023, 7, 1–4.
- 153 Z. Li, W. Zhang and F. Xing, Graphene Optical Biosensors, *Int. J. Mol. Sci.*, 2019, 20, 2461.
- 154 F. Ma, C. Li and C. Zhang, Development of quantum dot-based biosensors: principles and applications, *J. Mater. Chem. B*, 2018, 6, 6173–6190.
- 155 M. Shahrashoob, A. Mohsenifar, M. Tabatabaei, T. Rahmani-Cherati, M. Mobaraki, A. Mota and T. R. Shojaei, Detection of *Helicobacter Pylori* Genome with an Optical Biosensor Based on Hybridization of Urease Gene with a Gold Nanoparticles-Labeled Probe, *J. Appl. Spectrosc.*, 2016, 83, 322–329.
- 156 M. Rong, G. Kavokin and A. Sawan, Optical Biosensor Based on Porous Silicon and Tamm Plasmon Polariton for Detection of CagA Antigen of *Helicobacter Pylori*, *Sensors*, 2024, 24, 5153.
- 157 J. Tan, W. Geng, J. Li, Z. Wang, S. Zhu and X. Wang, Colorimetric and Fluorescence Dual-Mode Biosensors Based on Peroxidase-Like Activity of the Co₃O₄ Nanosheets, *Front. Chem.*, 2022, 10, 871013.
- 158 D. Zhu, B. Liu and G. Wei, Two-Dimensional Material-Based Colorimetric Biosensors: A Review, *Biosensors*, 2021, 11, 259.
- 159 G. A. R. Y. Suaifan, M. B. Shehadeh, R. M. Darwish, M. Alterify, W. Abu Jbara, F. Abu Jbara, N. Alaridah and M. Zourob, Magnetic beads-based nanozyme for portable colorimetric biosensing of *Helicobacter pylori*, *Biosens. Bioelectron. X*, 2024, 20, 100517.
- 160 Y. Fei, R. Fang, L. Xiao, Y. Zhang, K. Fan, Y. Jiang, S. Lei, R. Xu, D. Yang, Y. Ye, S. Xiang, P. Wang, C. Zhou and T. Tang, The development of a colorimetric biosensing assay for the detection of *Helicobacter pylori* in feces, *Anal. Biochem.*, 2022, 651, 114737.
- 161 Y. Kohli, T. Kato, S. Ito, H. Miyaji, T. Azuma, H. Tsuruta, S. Matsui, K. Oka and M. Nakamura, Endoscopic Urease Sensor System for Detecting *Helicobacter pylori* on Gastric Wall Using a pH-sensitive Field Effect Transistor, *Dig. Endosc.*, 1995, 7, 215–219.
- 162 D. Sadighbayan, M. Hasanzadeh and E. Ghafar-Zadeh, Biosensing based on field-effect transistors (FET): Recent progress and challenges, *TrAC Trends Anal. Chem.*, 2020, 133, 116067.
- 163 C.-C. Huang, Y.-H. Kuo, Y.-S. Chen, P.-C. Huang and G.-B. Lee, A miniaturized, DNA-FET biosensor-based microfluidic system for quantification of two breast cancer biomarkers, *Microfluid. Nanofluidics*, 2021, 25, 33.
- 164 L. Chao, H. Shi, K. Nie, B. Dong, J. Ding, M. Long and Z. Liu, Applications of Field Effect Transistor Biosensors Integrated in Microfluidic Chips, *Nanosci. Nanotechnol. Lett.*, 2020, 12, 427–445.
- 165 S. Lakshmi, S. Sundar, S. Menon, S. N. Kesana and M. G. Varshitha, 4th International Conference on Circuits, Control, Communication and Computing (I4C), IEEE, 2022, pp. 212–215.
- 166 G. Dyankov, T. Eftimov, E. Hikova, H. Najdenski, V. Kussovski, P. Genova-Kalou, V. Mankov, H. Kisov, P. Veselinov, S. S. Ghaffari, M. Kovacheva-Slavova, B. Vladimirov and N. Malinowski, SPR and Double Resonance LPG Biosensors for *Helicobacter pylori* BabA Antigen Detection, *Sensors*, 2024, 24, 2118.
- 167 P. Kolev, E. Hikova, H. Kisov and G. Dyankov, Biosensor-based serological assay for diagnosing *Helicobacter pylori* infection, *IOP Conf. Ser. Earth Environ. Sci.*, 2024, 1305, 012019.
- 168 H. Kim, H. Choi, Y. Heo, C. Kim, M. Kim and K. T. Kim, Biosensors Based on Bivalent and Multivalent Recognition by Nucleic Acid Scaffolds, *Appl. Sci.*, 2022, 12, 1717.
- 169 A. Sharma, K. Dulta, R. Nagraik, K. Dua, S. K. Singh, D. Kumar and D.-S. Shin, Potentialities of aptasensors in cancer diagnosis, *Mater. Lett.*, 2022, 308, 131240.
- 170 M. B. Kulkarni, N. H. Ayachit and T. M. Aminabhavi, Recent Advancements in Nanobiosensors: Current Trends, Challenges, Applications, and Future Scope, *Biosensors*, 2022, 12, 892.
- 171 M. Zhang, C. Liao, C. H. Mak, P. You, C. L. Mak and F. Yan, Highly sensitive glucose sensors based on enzyme-modified whole-graphene solution-gated transistors, *Sci. Rep.*, 2015, 5, 8311.
- 172 Biosensor Market Size & Share Analysis - Growth Trends & Forecasts (2024–2029) Source: <https://www.mordorintelligence.com/industry-reports/biosensor-market>.
- 173 T. Beduk, D. Beduk, M. R. Hasan, E. Guler Celik, J. Kosel, J. Narang, K. N. Salama and S. Timur, Smartphone-Based Multiplexed Biosensing Tools for Health Monitoring, *Biosensors*, 2022, 12, 583.
- 174 J. Yoon, J. Lim, M. Shin, S.-N. Lee and J.-W. Choi, Graphene/MoS₂ Nanohybrid for Biosensors, *Materials*, 2021, 14, 518.
- 175 X. Sun, J. Fan, C. Fu, L. Yao, S. Zhao, J. Wang and J. Xiao, WS₂ and MoS₂ biosensing platforms using peptides as probe biomolecules, *Sci. Rep.*, 2017, 7, 10290.
- 176 S. Afsahi, M. B. Lerner, J. M. Goldstein, J. Lee, X. Tang, D. A. Bagarozzi, D. Pan, L. Locascio, A. Walker, F. Barron and B. R. Goldsmith, Novel graphene-based biosensor for early detection of Zika virus infection, *Biosens. Bioelectron.*, 2018, 100, 85–88.
- 177 R. Li, Y. Wang, J. Du, X. Wang, A. Duan, R. Gao, J. Liu and B. Li, Graphene oxide loaded with tumor-targeted peptide and anti-cancer drugs for cancer target therapy, *Sci. Rep.*, 2021, 11, 1725.
- 178 P. Basu, A. Banerjee, P. D. Okoro, A. Masoumi, B. Kanjilal, M. Akbari, M. Martins-Green, D. G. Armstrong and I. Noshadi, Integration of Functional Polymers and Biosensors to Enhance Wound Healing, *Adv. Healthcare Mater.*, 2024, 13, 2401461.
- 179 S. Shrestha, B. Wang and P. Dutta, Nanoparticle processing: Understanding and controlling aggregation, *Adv. Colloid Interface Sci.*, 2020, 279, 102162.
- 180 H. Rasuli and R. Rasuli, Nanoparticle-decorated graphene/graphene oxide: synthesis, properties and applications, *J. Mater. Sci.*, 2023, 58, 2971–2992.



- 181 J. Wu, H. Lin, D. J. Moss, K. P. Loh and B. Jia, Graphene oxide for photonics, electronics and optoelectronics, *Nat. Rev. Chem.*, 2023, 7, 162–183.
- 182 S. Zhang and C. Wang, Precise Analysis of Nanoparticle Size Distribution in TEM Image, *Methods Protoc.*, 2023, 6, 63.
- 183 J. Wang, K. Fu, X. Zhang, Q. Yin, G. Wei and Z. Su, When MoS₂ meets TiO₂: facile synthesis strategies, hybrid nanostructures, synergistic properties, and photocatalytic applications, *J. Mater. Chem. C*, 2021, 9, 8466–8482.
- 184 Y. Wu, J.-Y. Chen and W.-M. He, Surface-enhanced Raman spectroscopy biosensor based on silver nanoparticles@metal-organic frameworks with peroxidase-mimicking activities for ultrasensitive monitoring of blood cholesterol, *Sensors Actuators B Chem.*, 2022, 365, 131939.
- 185 Y. F. Duan, Y. Ning, Y. Song and L. Deng, Fluorescent aptasensor for the determination of Salmonella typhimurium based on a graphene oxide platform, *Microchim. Acta*, 2014, 181, 647–653.
- 186 Y. Xie, X. Zhi, H. Su, K. Wang, Z. Yan, N. He, J. Zhang, D. Chen and D. Cui, A Novel Electrochemical Microfluidic Chip Combined with Multiple Biomarkers for Early Diagnosis of Gastric Cancer, *Nanoscale Res. Lett.*, 2015, 10, 477.
- 187 R. Salimian, S. Shahrokhian and S. Panahi, Enhanced Electrochemical Activity of a Hollow Carbon Sphere/Polyaniline-Based Electrochemical Biosensor for HBV DNA Marker Detection, *ACS Biomater. Sci. Eng.*, 2019, 5, 2587–2594.
- 188 Z. Qian, X. Shan, L. Chai, J. Chen and H. Feng, Simultaneous Detection of Multiple DNA Targets by Integrating Dual-Color Graphene Quantum Dot Nanoprobes and Carbon Nanotubes, *Chem. – A Eur. J.*, 2014, 20, 16065–16069.
- 189 S. He, B. Song, D. Li, C. Zhu, W. Qi, Y. Wen, L. Wang, S. Song, H. Fang and C. Fan, A Graphene Nanoprobe for Rapid, Sensitive, and Multicolor Fluorescent DNA Analysis, *Adv. Funct. Mater.*, 2010, 20, 453–459.
- 190 P. Labroo and Y. Cui, Graphene nano-ink biosensor arrays on a microfluidic paper for multiplexed detection of metabolites, *Anal. Chim. Acta*, 2014, 813, 90–96.
- 191 F. Mi, C. Hu, Y. Wang, L. Wang, F. Peng, P. Geng and M. Guan, Recent advancements in microfluidic chip biosensor detection of foodborne pathogenic bacteria: a review, *Anal. Bioanal. Chem.*, 2022, 414, 2883–2902.
- 192 M. B. Kulkarni, N. H. Ayachit and T. M. Aminabhavi, Recent Advances in Microfluidics-Based Electrochemical Sensors for Foodborne Pathogen Detection, *Biosensors*, 2023, 13, 246.
- 193 S. Chen, Z. Qiao, Y. Niu, J. C. Yeo, Y. Liu, J. Qi, S. Fan, X. Liu, J. Y. Lee and C. T. Lim, Wearable flexible microfluidic sensing technologies, *Nat. Rev. Bioeng.*, 2023, 1, 950–971.
- 194 M. Gray, J. Meehan, C. Ward, S. P. Langdon, I. H. Kunkler, A. Murray and D. Argyle, Implantable biosensors and their contribution to the future of precision medicine, *Vet. J.*, 2018, 239, 21–29.
- 195 X. Mei, D. Ye, F. Zhang and C. Di, Implantable application of polymer-based biosensors, *J. Polym. Sci.*, 2022, 60, 328–347.
- 196 P. Puig-Serra, M. C. Casado-Rosas, M. Martinez-Lage, B. Olalla-Sastre, A. Alonso-Yanez, R. Torres-Ruiz and S. Rodriguez-Perales, CRISPR Approaches for the Diagnosis of Human Diseases, *Int. J. Mol. Sci.*, 2022, 23, 1757.
- 197 X. Zhang, H. Qiu, X. Zhong, S. Yi, Z. Jia, L. Chen and S. Hu, A CRISPR/Cas12a-assisted array for Helicobacter pylori DNA analysis in saliva, *Anal. Chim. Acta*, 2023, 1239, 340736.
- 198 E. Qiu, S. Jin, Z. Xiao, Q. Chen, Q. Wang, H. Liu, C. Xie, C. Chen, Z. Li and S. Han, CRISPR-based detection of Helicobacter pylori in stool samples, *Helicobacter*, 2021, 26, 12828.
- 199 H. Liu, J. Wang, X. Hu, X. Tang and C. Zhang, A rapid and high-throughput Helicobacter pylori RPA-CRISPR/Cas12a-based nucleic acid detection system, *Clin. Chim. Acta*, 2023, 540, 117201.
- 200 F. Cui, Y. Yue, Y. Zhang, Z. Zhang and H. S. Zhou, Advancing Biosensors with Machine Learning, *ACS Sens.*, 2020, 5, 3346–3364.
- 201 S. Kadian, P. Kumari, S. Shukla and R. Narayan, Recent advancements in machine learning enabled portable and wearable biosensors, *Talanta Open*, 2023, 8, 100267.
- 202 P. P. Das, R. Nagraik, A. Sharma, T. K. Upadhyay, H. Lahlhennawia, D. Balram, K.-Y. Lian, J. Singh and D. Kumar, recent update on biomimetic sensor technology for cancer diagnosis, *Talanta Open*, 2024, 9, 100276.
- 203 R. Verma, S. K. Yadav, R. B. K. Singh, R. Verma, D. Kumar and J. Singh, Laccase-Conjugated Nanostructured ZnFe₂O₄/rGO-Modified Electrode-Based Interfaces for Electrochemical Impedance Monitoring of Adrenaline: A Promising Biosensor for Management of Neurodegenerative Disorders, *ACS Appl. Bio Mater.*, 2023, 6(12), 5842–5853.
- 204 S. S. Chu, H. A. Nguyen, J. Zhang, S. Tabassum and H. Cao, Towards Multiplexed and Multimodal Biosensor Platforms in Real-Time Monitoring of Metabolic Disorders, *Sensors*, 2022, 14, 5200.
- 205 R. Tomichan, A. Sharma, K. Akash, A. A. Siddiqui, A. Dubey, T. K. Upadhyay, D. Kumar, S. Pandey and R. Nagraik, Insight of smart biosensors for COVID-19: A review, *Luminescence*, 2023, 38, 1102.
- 206 B. Gil Rosa, O. E. Akingbade, X. Guo, L. Gonzalez-Macia, M. A. Crone, L. P. Cameron, P. Freemont, K.-L. Choy, F. Güder, E. Yeatman, D. J. Sharp and B. Li, Multiplexed immunosensors for point-of-care diagnostic applications, *Biosens. Bioelectron.*, 2022, 203, 114050.

